NIST Calibration Program Calibration Services Users Guide SP 250 Appendix Fee Schedule 2006

Calibration Services:

Dimensional
Mechanical
Thermodynamic
Optical Radiation
Ionizing Radiation
Electromagnetic
Time and Frequency

National Institute of Standards and Technology Technology Administration U.S. Department of Commerce

Table of Contents

CHAPTER 1		
POLIC		
	Introduction	4
	Types of Services	
	Other NIST Measurement Transfer Services	
	Criteria for Quality Assurance	
	Fees	
	Reports of Calibration/Test Results	
	Traceability	
	NIST Policy on Reporting Measurement Uncertainty	
	NIST Policy Regarding Use of Metric (SI) Units	
	References to NIST in Advertisements	
	DisclaimerQuestions and Inquires	
	Questions and inquires	
CHAPTER 2	2	
ORDE	RING INSTRUCTIONS FOR DOMESTIC CUSTOMERS	
	Customer Inquires	9
	Prearrangement and Scheduling	9
	Purchase Orders	9
	Shipping, Insurance, and Risk of Loss	
	Turnaround Time	10
	Customer Checklist	10
CHAPTER 3 SPECI	AL INSTRUCTIONS FOR FOREIGN CUSTOMERS Foreign Inquires Criteria for Providing Service Special Instructions Shipping Charges	12 12
CHAPTER 4	NSIONAL MEASUREMENTS Length Measurements	
	Diameter Measurements	
	Complex Dimensional Standards	
	Optical Reference Planes and Roundness Standards	
	Angular Measurements	
	Surface Texture	
CHAPTER 5	5 IANICAL MEASUREMENTS	
	Hydrometers	
	Volume and Density	
	Flow Measurements	
	Flow Measurements at Cryogenic Temperatures	
	Air Speed Instruments	
	Mass Standards	
	Force Measurements	26

Vibration Me	easurements	27
Acoustic Mea	asurements	27
CHAPTER 6		
THERMODYNAMIC	QUANTITIES	
	asurements	29
Low Pressure	e, Vacuum and Leak Measurements	29
Laboratory a	nd Industrial-Grade Thermometers	31
	les, Thermocouple Materials, Thermometer Indicators	
	hermometry	
	mperature Measurements	
Humidity Me	easurements	37
CHAPTER 7		
OPTICAL RADIATION	ON MEASUREMENTS	
	Measurements	
Optical Prope	erties of Materials Measurements	39
Spectroradion	metric Measurements	39
	Standards in the Ultraviolet	
Laser and Op	otoelectronic Components Used with Lasers	42
CHAPTER 8		
IONIZING RADIATI	ON MEASUREMENTS	
Radioactivity	Sources	44
	rces and Neutron Dosimetry	
Dosimetry of	X-Rays, Gamma-Rays	45
Dosimetry fo	or High-Dose Applications	46
CHAPTER 9		
ELECTROMAGNET	IC MEASUREMENTS	
Resistance M	leasurements	48
	Measurements (Except Resistors)	
	surements	
Precision Rat	tio Measurements	55
	and Standards and VOR Measurements	
	nergy Measurements, Low-Frequency	
	ive and Millimeter-Wave Measurements	
	etic Field Strength and Antenna Measurements	
Pulse Wavefo	orm Measurements	66
CHAPTER 10		
	ENCY MEASUREMENTS	
	d Measurement Services	
	nd Characterization of Oscillators and Amplifiers	
Tests of PM/A	AM Noise Measurement Systems	68

CHAPTER 11	
SEMINARS	
NIST Measurement Seminars	70
ALPHABETICAL CROSS-INDEX	71

CHAPTER 1 POLICIES

A. Introduction

The calibration services of the National Institute of Standards and Technology (NIST) are designed to help the makers and users of precision instruments achieve the highest possible levels of measurement quality and productivity. The services listed in this Fee Schedule constitute the highest order of calibration services available in the United States. They directly link a customer's precision equipment or transfer standards to national and international measurement standards. These services are offered to public and private organizations and individuals alike.

For more specific information, the NIST Calibration Services Users Guide, SP 250, contains data on uncertainty and other technical references. Copies are available upon request or consult our web site (see Section L).

B. Types of Calibration Services

- Calibration Services
- Special Tests
- Measurement Assurance Programs (MAPs)

NIST provides Calibration Services using well-characterized, stable and predictable measurement processes. NIST calibrates instruments and devices that are metrologically suitable as reference or transfer standards.

Special Tests are either unique or seldom-performed calibrations or measurements requested by the customer.

Measurement Assurance Programs are quality control programs for calibrating a customer's entire measurement system. In a typical MAP, a stable artifact or set of artifacts called transfer standards are first measured by NIST and then sent to a customer's laboratory for a series of measurements. The transfer standards are then returned to NIST for remeasurement, along with the participating laboratory's results. NIST reports its comparative findings to the customer and, when necessary, offers guidance on achieving and maintaining measurement quality. Successful use of a NIST MAP requires that the customer make periodic measurements of in-house check standards to estimate their measurement process uncertainty and to ensure that the measurement process remains in a state of statistical control. Unless a laboratory has a measurement quality assurance program to monitor its own measurement process parameters continuously, there is no value in participating in a MAP. In fact, NIST recommends that its customers establish and use a measurement quality assurance program to monitor their measurement parameters, whether or not they participate in a MAP.

C. Other NIST Measurement Transfer Services

National Voluntary Laboratory Accreditation Program (NVLAP)

NIST does not audit or regulate metrology laboratories as part of MAP or other calibration services. Calibration laboratories and testing facilities may be accredited by NIST under the National Voluntary Accreditation Program (NVLAP). The basic procedures and general accreditation requirements of NVLAP are described in NIST Handbook 150. A participating laboratory may voluntarily take steps to improve or assess its measurement process. For further information about NVLAP, contact:

National Voluntary Laboratory Accreditation Program (NVLAP)

National Institute of Standards and Technology

Telephone: (301) 975-4016

Fax: (301) 926-2884

100 Bureau Drive, Stop 2140

Email: NVLAP@nist.gov

Gaithersburg, MD 20899-2140

Internet: www.ts.nist.gov/nvlap

Standard Reference Materials Group (SRM)

Calibration assistance and alternative paths for traceability are provided by NIST's Standard Reference Materials Group. Chemical measurement instruments are not calibrated at NIST, but NIST provides suites of Standard Reference Materials (SRMs) for the calibration of the instrument by the user. In addition, NIST provides SRMs for dimensional measurements, thermodynamic property and photometric measurements. For further information about SRMs, contact:

Standard Reference Materials Group (SRM)

National Institute of Standards and Technology

100 Bureau Drive, Stop 2322

Gaithersburg, MD 20899-2322

Standard Reference Materials Group (SRM)

Fax: (301) 948-3730

Email: srminfo@nist.gov

Internet: www.nist.gov/srm

Standard Reference Data Group (SRD)

Very few calibrations can be conducted without additional quantitative information related to measurement of physical or chemical properties. NIST develops and publishes evaluated data for technical and scientific applications called Standard Reference Data. For further information about SRD, contact:

Standard Reference Data Group (SRD)

National Institute of Standards and Technology

Fax: (301) 975-2208

Fax: (301) 926-0416

100 Bureau Drive, Stop 2310

Gaithersburg, MD 20899-2310

Internet: www.nist.gov/srd

Weights and Measures Division (W&M)

The NIST Weights and Measures Division (W&M) provides measurement services to State and local governments responsible for marketplace transactions involving measurements. State weights and measures laboratories provide alternative sources for calibration services in mass, length, volume, and certain other measurement areas. For further information contact:

Weights and Measures Division (W&M)Telephone:(301) 975-4004National Institute of Standards and TechnologyFax:(301) 926-0647100 Bureau Drive, Stop 2600Email:owm@nist.govGaithersburg, MD 20899-2600Internet:www.nist.gov/owm

D. Criteria for Quality Assurance

All the measurement services listed in this document meet rigorous criteria for quality assurance. Calibration Services and MAPs satisfy the most demanding and explicit requirements in that they are carried out regularly under pre-established and well-defined conditions; the measurement processes involved are well-characterized, stable, and statistically controlled; and quality-control procedures are well-defined and strictly followed. Furthermore, each Calibration Service or MAP is planned and documented to permit continuity of service over time.

A Special Test is so designated for one or more of the following reasons: (1) the specific type of calibration is seldom requested, thus precluding the maintenance of a large statistical base for characterizing the measurement process; (2) the test request is unique; or (3) the service is still under development—meaning the measurement or calibration methods are still being perfected, or all quality-control documentation not been completed.

E. Fees

NIST recovers the cost of providing calibration services by charging a fee for each calibration performed. The costs of services are published in the Fee Schedule, which is updated and published annually to reflect changes in prices and services. Even so, the cost of many services varies according to your exact calibration specifications; you must therefore provide the technical contact with an exact description of work before receiving a price quote.

NOTE: Fees for NIST services do not include shipping costs or insurance.

F. Reports of Calibration/Test Results

Reports on calibrations or other services are the property of the customer. Copies are supplied to other parties only as required by federal law or requested in writing by the customer. The results of calibrations and tests performed by NIST apply only to the specific instrument or standard at the time of test unless otherwise clearly stated.

G. Traceability

The International Vocabulary of Basic and General Terms in Metrology (VIM; 1993) defines traceability as:

The property of the result of measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties.

Many government regulations and commercial contracts require regulated organizations or contractors to verify that the measurements they are "traceable" and to support the claim of traceability by keeping records that their own measuring equipment has been calibrated by laboratories or testing facilities whose measurements are part of this "unbroken chain." The purpose of requiring traceability is to ensure that measurements are accurate representations of the specific quantity subject to measurement, within the uncertainty of the measurement.

NIST reports its calibration results, with the measurement values accompanied by the uncertainties associated with the methods, operators, and environment at NIST. Users of these calibration services will make their own measurements with the calibrated instruments or artifacts. In addition to the uncertainty indicated by NIST, other uncertainties are inherent in the instrument, associated with the method or protocol in using the instrument, with the operator of the instrument, and with the physical environment (pressure, temperature, humidity, etc.) in which the measurements are made. Thus, the measurements made with the calibrated instruments or artifacts by organizations outside of NIST have total uncertainty budgets associated with them, only one component of which is the uncertainty reported to them by NIST.

NIST often receives calls to verify the authenticity of a NIST Report of Test number appearing on another organization's report. Although NIST can verify the authenticity of its report numbers, having an authentic number does not provide complete assurance or evidence that the measurement value provided by another organization is traceable. Not only should there be an unbroken chain of comparisons, each provided measurement should be accompanied by a statement of uncertainty associated with farthest link in the chain from NIST, that is, the last facility providing the measurement value. NIST does not have that information; only the facilities that provided the measurement values to the customer can provide the associated uncertainties and describe the traceability chain.

In summary, to adequately establish an audit trail for traceability, a proper calibration result should include: the assigned value, a stated uncertainty, identification of the standards used in the calibration, and the specification of any environmental conditions of the calibration where correction factors should be applied, if the standard or equipment were to be used under different environmental conditions.

NIST does not define nor enforce traceability except in its NVLAP laboratory accreditation program. Moreover, NIST is not legally required to comply with traceability requirements of other federal agencies; nor do we determine what must be done to comply with another party's contract or regulation calling for such traceability. However, NIST can and does provide technical advice on making measurements consistent with national standards.

Although NIST supports making the user aware of traceability and provides the user with details as to how traceability is established, NIST does not allow the prominent display of its name on proprietary products or in the advertising of them. (See Section J).

H. NIST Policy on Reporting Measurement Uncertainty

To ensure that NIST uncertainty statements are consistent across the organization and with international practice, NIST policy requires that all NIST measurements be accompanied by statements of uncertainty as discussed in NIST Technical Note 1297¹. That publication is based on the approach to expressing uncertainty in measurements recommended by the International Committee on Weights and Measures (CIPM)². That committee established general rules for evaluating and expressing uncertainty in measurements that are intended to be applicable to a broad spectrum of measurements. Copies of NIST TN 1297 are available upon request (see Section L) or on the web site: www.physics.nist.gov/Pubs/guidelines/contents.html.

The American National Standard for Expressing Uncertainty-U.S. Guide to the Expression of Uncertainty of Measurement (ANSI/NCSL Z540-2-1997) is available from the

NCSL International 2995 Wilderness Place, Suite 107 Boulder, CO 80301-5404 303/440-3339 www.ncsli.org

NIST reports its calibration results with the measurement values accompanied by the uncertainties associated with the methods, operators, and environment at NIST. Users of these calibration services will make their own measurements with the calibrated instruments or artifacts. In addition to the uncertainty indicated by NIST, other uncertainties are inherent in the instrument, associated with the method or protocol in using the instrument, with the operator of the instrument, and with the physical environment (pressure, temperature, humidity, etc.) in which the measurements are made. Thus, the measurements made with the calibrated instruments or artifacts by organizations outside of NIST have total uncertainty budgets associated with them, only one component of which is the uncertainty reported to them by NIST.

I. NIST Policy Regarding Use of Metric (SI) Units

In accordance with the Metric Conversion Act of 1975 as amended by Section 5164 of the Omnibus Trade and Competitiveness Act of 1988 and as required by related provisions of the Code of Federal Regulations, the National Institute of Standards and Technology (NIST) uses the modern metric system of measurement units (International System of Units–SI) in all publications. When the field of application or the special needs of users of NIST publication require the use of non-SI units, the values of quantities are first stated in the SI units and the corresponding values expressed in non-SI units follow in parentheses. Copies of NIST SP 811³ are available upon request (see Section L) or on the web site: www.physics.nist.gov/Pubs/SP811/sp811.html

¹Guideline for Evaluating and Expressing the Uncertainty of NIST Measurement Results, NIST Technical Note 1297, 1994 Edition.

²Guide to the Expression of Uncertainty in Measurement, International Standards Organization (ISO), 1993 Edition.

³Guide for the Use of the International System of Units (SI), NIST Special Publication 811, 1995 Edition.

J. Reference to NIST in Advertisements

The NIST measurement/test results or reports shall not be used to indicate or imply that NIST approves, recommends, or endorses the manufacturer, supplier, or user of any instruments or standards or that NIST in any way guarantees or predicts the future performance of items after calibration or test. No reference shall be made to NIST or to reports or results furnished by NIST in any advertising or sales promotions, which would indicate or imply that NIST approves, recommends, or endorses any proprietary product or proprietary material.

K. Disclaimer

Commercial products, materials, and instruments, are identified in our communications and documents for the sole purpose of adequately describing experimental or test procedures. In no event does such identification imply recommendation or endorsement by NIST of a particular product; nor does it imply that a named material or instrument is necessarily the best available for the purpose it serves.

L. Questions and Inquires

This Fee Schedule is intended to make the task of selecting and ordering an appropriate calibration service as quick and easy as possible. Nevertheless, when questions arise, you should contact NIST for immediate clarification.

General inquires about the NIST calibration services, assistance in determining the availability of services, and requests for complimentary copies of the Calibration Services Users Guide, the Fee Schedule, Guide for the International System of Units (SP 811), and Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results (TN 1297) are to be addressed to:

Calibration Program
National Institute of Standards and Technology
100 Bureau Drive, Stop 2330
Gaithersburg, MD 20899-2330

Telephone: (301) 975-2092
Fax: (301) 869-3548
Email: calibrations@nist.gov
Internet: www.ts.nist.gov/calibrations

For technical questions concerning a specific service, directly contact the NIST staff member responsible for that calibration area.

CHAPTER 2

ORDERING INSTRUCTIONS FOR DOMESTIC CUSTOMERS

A. Customer Inquires

General customer inquires for information or clarifications about the NIST calibration services may be directed as indicated in Section L of Chapter 1.

B. Prearrangements and Scheduling

Services should be arranged in advance, beginning with direct contact with a NIST technical staff member responsible for the desired service. Use the appropriate technical section of the Users Guide or Fee Schedule to determine whom to contact. This advance communication may answer your questions, clarify the policies and procedures briefly described here, and will permit you to schedule a tentative calibration date. Following the initial communication, you must complete and submit a purchase order and prepare to ship the item according to the procedures described below or agreed upon with the technical contact. If a calibration is scheduled far in advance, the item should not be shipped until shortly before the scheduled date; you must submit the purchase order (complete with the name and number of the desired service) before a firm calibration date can be assigned. When NIST receives your purchase order and assigns a firm service date, your order will be confirmed by the technical contact.

C. Purchase Orders

Before you ship an item for calibration, send a purchase order to the address listed in the appropriate technical section of the Users Guide or Fee Schedule. The purchase order must:

- 1. State both the name and number of the NIST service (listed in this Fee Schedule as the "Service ID Number") being requested. FAILURE TO INCLUDE THE SERVICE ORDER NUMBER WILL SERIOUSLY IMPEDE SCHEDULING AND SERVICE.
- 2. Clearly identify the item(s) being send for calibration, including any serial number(s) or model number(s).
- 3. Give the name, address, and telephone number of your company's procurement officer, purchasing agent or other administrative/financial authority.
- 4. Give the name, address, and telephone number of your company's technical contact, if different from above.
- 5. List separately the instructions and address for return shipment, insurance, mailing address for the calibration/test report, and billing address. (Federal or state agency requests for calibration services should be accompanied by a document authorizing that the cost of the service be billed to the agency.)
- 6. Clearly state any special or necessary conditions of test, such as operating frequency or temperature.
- 7. Clearly state the customer identification number; i.e., social security number (EIN) for individuals; tax identification number (TIN) for organizations; or agency location code (ALC) for government customers.
- 8. If the calibration or test report is to be handled in a special manner, give instructions on the purchase order.

NOTE: Receipt of orders by NIST does not imply acceptance of any provisions set forth in the order that are contrary to the policy, practice, or regulations of NIST or the U.S. Government. In general, NIST will not sign any affidavits, acknowledgement forms, or other documents that may be required by company policy governing the procurement of goods and services.

D. Shipping, Insurance, and Risk of Loss

Ship the instrument or standard to the mailing address of the technical group providing the service. Please take note that the mailing address is not the same for every technical group.

Please adhere rigorously to the following procedures:

- 1. Ship only items in good repair. Apparatus in disrepair will not be calibrated. If defects are found after calibration has begun, the procedure will be terminated, a report issued, and a charge levied for work completed.
- Use strong, reusable packing materials and containers marked clearly and indelibly on the
 outside with the requestor's name, address and the following notation: REUSABLE
 CONTAINER, DO NOT DESTROY.
- Follow any special shipping procedures given in the technical sections of the Calibration Services Users Guide, particularly those sections covering radiation and dosimetry measurements.
- 4. Insure the shipments to and from NIST and clearly state the method of return shipment. NIST will not assume liability for loss or damage unless such loss and damage result solely from the negligence of NIST personnel. If return shipment by parcel post is requested or is suitable, NIST will prepay the return shipment but will not insure it. When no shipping or insurance instructions are furnished, NIST will return the shipment by common carrier, collect and uninsured.
- 5. Shipments to NIST must be at FOB destinations (customer pays for shipping.)
- 6. Return shipments are sent FOB origin (customer pays for shipping.)

NOTE: Fees for NIST services do not include shipping cost or insurance.

E. Turnaround Time

Normal turnaround time for NIST calibration services varies greatly—usually from several weeks to several months depending on the type of service requested, and the service schedule. Some services are only scheduled once or twice a year with appointments made months in advance of the service date. To avoid unnecessary scheduling or administrative delays in the calibration process, always make arrangements with the technical contact for the service you wish to utilize prior to shipping your instrument or artifact to us.

F. Customer Checklist

Please refer to page 11 in this chapter for a Customer Checklist which is intended to assist you in developing the basic information required to process an order for calibration services at NIST.

Customer Checklist for Ordering NIST Calibration Services

Information Obtained from NIST Technical Contact	Comments
NIST Contact (name/telephone)	Provide this information on your purchase order (po)
Is the service available?	Please make sure customer's technical contact discusses service with NIST technical contact before proceeding.
NIST Service Identification Number	Provide this information on your po
Estimated cost of services	Provide this information on your po
Estimated turnaround time	Many calibration services are batched. Find out when to send the instrument.
Special instructions	
Packaging instructions	
Shipping instructions	
Other Precautions	
Information Supplies by the Customer on Purchase Order	
Purchase order number	
Purchase order date	
Customer's tax identification number	
Customer's mailing address	
Customer's billing address	
Name, telephone number, fax number, email address of administrative or procurement contact point at customer's location	
Name, telephone number, fax number, email address of technical contact point at customer's location	
Ship-to address (including NIST technical contact name)	
Return address (for shipment back to customer)	
NIST Service Identification Number	
Estimated cost	
Shipping terms (no FOB destination on return shipment)	
Special instructions from customer's technical contact	

CHAPTER 3

SPECIAL INSTRUCTIONS FOR FOREIGN CUSTOMERS

A. Foreign Inquires

Foreign customers should address all inquiries to:

Calibration Program
National Institute of Standards and Technology
100 Bureau Drive, Stop 2330
Gaithersburg, MD 20899-2330
United States of America

Telephone: (301) 975-2092
Fax: (301) 869-3548
Email: calibrations@nist.gov
Internet: www.ts.nist.gov/calibrations

NOTE: Please clearly indicate your **city** and **country** on all correspondence so that we may promptly respond to your request.

B. Criteria for Providing Service

Under certain circumstances, NIST is authorized to provide measurement service, including calibration services, for organizations or individuals located outside the United States. However, the Calibration Program must review each request for calibration services to determine if services are available to the requestor's organization in the requestor's country. Foreign customers must provide the following information, in writing, to the Calibration Program (see address above):

- 1. Identification of the item(s) to be calibrated, including serial and model numbers.
- A detailed description of the measurements that are needed, or indicate the service identification number.
- 3. A description of any special requirement/circumstance that might affect the decision to provide the service. For example, will adjustments have to be made to the instrument, or will the time period be restricted in which the device is available for calibration?
- 4. A complete name and address of the requestor's organization.

C. Special Instructions

If the request for calibration service is accepted by NIST, the requesting organization will be notified of the cost of service and will be given the contact information for the NIST technical unit that will perform the measurements. The requesting organization must then complete the following steps:

- 1. Contact the NIST technical staff that will perform the service to determine the time schedule.
- 2. Send a purchase order to the Calibration Program. Provide complete addresses, including country, for returning the instrument and for mailing the calibration or test report.
- 3. NIST policy requires prepayment for all NIST calibration services requested by non-U.S. organizations. Before proceeding with any service(s) we will need a check, money order or a bank wire transfer. The prepayment must be for the full amount and be drawn on a U.S. bank. The prepayment methods are as follows:

Money Orders & Prepayment Check

Checks made payable to the National Institute of Standards and Technology (NIST) should be mailed to:

Calibration Program
National Institute of Standards and Technology
100 Bureau Drive, Stop 2330
Gaithersburg, MD 20899-2330

Bank Wire Transfers

Treas NYC (Account is with the Federal Reserve Bank of New York) U.S. Dept. of Treasury FMS-Banking Operations Branch 3700 East West Highway, Room 5A05 Hyattsville, MD 20782

Phone: 001 (202) 874-6617

ABA# 021030004 Account # 13060001

Account Name: TREAS NYC/CTR/BNF=/NIST/AC-13060001

Reference "Calibrations" to enable us to identify your payment. In addition, please be sure to pay any fees assessed for your bank wire transfers; otherwise, they will deduct it from your prepayment wire.

PLEASE NOTE: Our account number and name are of critical importance and must be referenced in order for NIST to be properly credited with your payment. It must appear in the precise manner shown to allow for the automated processing and classification of the funds transfer message. In addition, please refer to the NIST invoice number, your purchase order number, your country, and any other pertinent information that would help us identify you payment.

This transfer of funds can only be accomplished by your company going through a U.S. correspondent bank or by having your country's central bank send a swift telecommunication system message to the Federal Reserve Bank. Be sure to cover any processing fees your bank may charge you. Questions on bank wiring can be directed to the NIST Accounts Receivable Office at (301) 975-3880, email: billing@nist.gov, or fax at (301) 975-8943..

4. Before shipping the instrument or standard to the appropriate NIST technical unit, you must arrange with a customs broker for entry of the instrument into the U.S. with transportation to and from the port of entry to NIST prepaid. Air freight is most satisfactory. Entry bond is required for instruments not manufactured in the U.S. If arrangements are made with a broker in the country of origin, that broker should, in turn, have a U.S. customs broker in or near the port of entry to arrange for the entry of the instrument and its transportation to NIST. Direct arrangements can be made with customs brokers located in the Washington, DC/Baltimore, Maryland, metropolitan area or in the Denver, Colorado, area, as appropriate. These brokers must arrange for transportation to the port of exit after testing/calibration is completed.

D. Shipping Charges

The calibration costs quoted *do not* include shipping, insurance, or the services of a customs broker. You must arrange and pay for these services separately. For your information, NIST currently uses the following customs brokers:

Gaithersburg, MarylandBouldLaing InternationalFedEP.O. Box 161444725Washington, DC 20041DenvePhone: (703) 471-9279Phor

Phone: (703) 471-9279 Fax: (703) 471-8436 Boulder, Colorado

FedEx Trade Networks 4725 Paris Street, Suite 200

Denver, CO 80239 Phone: (303) 371-9550 Fax: (303) 373-0850

You are not required to use these customs brokers, but may select a broker of your choice.

CHAPTER 4

DIMENSIONAL MEASUREMENTS

A. Length Measurements A.1 Gage Blocks

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Stacey Adams (301) 975-3471 stacey.adams@nist.gov NIST

Theodore Doiron (301) 975-3472 theodore.doiron@nist.gov 100 Bureau Drive, Stop 8211 Gaithersburg, MD 20899-8211

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

A.1 Gage Blocks			
Service ID Number	Description of Services	Fee (\$)	
10010C	Gage Blocks: Set Up Charge, per order	150	
10011C	Mechanical Comparisons, per Block (100 mm and shorter)	93	
10012C	Mechanical Comparisons, per Block (over 100 mm)	217	
10013C	Interferometry, per Block (100 mm and shorter), Maximum 25 Blocks per Order	258	
10014C	Interferometry, per Block (over 100 mm), by Special Arrangement	At Cost	
10015C	Non-standard size Gage Blocks, by Special Arrangement	At Cost	

Fees are subject to change without notice.

A.2 Line Standards

Technical Contact:	Telephone:	<u>Email:</u>	Mailing Address:
William B. Penzes	(301) 975-3477	wpenzes@nist.gov	NIST
Thomas W. LeBrun	(301) 975-4256	thomas.lebrun@nist.gov	100 Bureau Drive, Stop 8212
			Gaithersburg, MD 20899-8212

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

A.2 Line Standards			
Service ID Number	Description of Service	Fee (\$)	
10020C	Line Standards: Scales, < 1 m (40 inches), 4 Passes	7328	
10021C	Line Standards: Scales, < 1 m (40 inches), 8 Passes	11329	
10022C	Line Standards: Stage Micrometer, Per Scale, 30 Intervals, 2 Passes	1227	
10023C	Line Standards: Stage Micrometer, Per Scale, 30 Intervals, 4 Passes	1548	
10024C	Line Standards: End Standards, < 1 m	7357	
10025C	Line Standards: Grid Plates, Less than 60 Intervals, 1 D Linear Calibration	12263	

A.3 Metal Tapes/Scales and Long Length Artifacts

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Ronald G. Hartsock (301) 975-3465 rhartsock@nist.gov NIST

100 Bureau Drive, Stop 8211 Gaithersburg, MD 20899-8211

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

	A.3 Metal Tapes/Scales and Long Length Artifacts	
Service ID Number	Description of Services	Fee (\$)
10030C	Metal Tapes: Surveying, Oil Gaging and General Purpose; Metal Scales	At Cost
10040S	Special Tests of Long Length Artifacts	At Cost

A.4 Other Length Standards

Technical Contacts: Telephone: Email: Mailing Address:

Eric S. Stanfield (301) 975-4882 eric.stanfield@nist.gov NIST

Theodore Doiron (301) 975-3472 theodore.doiron@nist.gov 100 Bureau Drive, Stop 8211

Gaithersburg, MD 20899-8211

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

	A.4 Other Length Standards	
Service ID Number	Description of Services	Fee (\$)
10050S	Special Tests of Length Standards	At Cost

A.5 Sieves

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Stacey Adams (301) 975-3471 stacey.adams@nist.gov NIST

Theodore Doiron (301) 975-3472 theodore.doiron@nist.gov 100 Bureau Drive, Stop 8211

Gaithersburg, MD 20899-8211

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

	A.5 Sieves	
Service ID Number	Description of Services	Fee (\$)
10060S	Special Test of Sieves	177

A.6 Algorithms Testing and Evaluation Program for Coordinate Measuring Systems

<u>Technical Contact:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Craig M. Shakarji (301) 975-3545 shakarji@nist.gov NIST

100 Bureau Drive, Stop 8260 Gaithersburg, MD 20899-8260

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

A.6 Algorithms Testing and Evaluation Program for Coordinate Measuring Systems			
Service ID Number	Description of Services	Fee (\$)	
10070S	Special Test of CMS Software: NIST-generated data sets (basic service)	1437	
10071S	Special Test of CMS Software: NIST-generated data sets (per geometry evaluated)	At Cost	
10072S	Special Test of CMS Software: NIST-generated data sets, standard level (per geometry evaluated)	733	
10080S	Special Test of CMS Software: Customer-generated data sets (basic service)	1898	
10081S	Special Test of CMS Software: Customer-generated data sets (per geometry evaluated)	At Cost	
10082S	Special Test of CMS Software: Customer-generated data sets, standard level (per geometry evaluated)	785	

Fees are subject to change without notice.

B. Diameter Measurements

<u>Technical Contacts:</u>	<u>Telephone:</u>	<u>Email:</u>	Mailing Address:
Eric S. Stanfield	(301) 975-4882	eric.stanfield@nist.gov	NIST
(11010S-11060S)			100 Bureau Drive, Stop 8211
Stacey Adams	(301) 975-3471	stacey.adams@nist.gov	Gaithersburg, MD 20899-8211
(11020C-11021C)			
TheodoreDoiron	(301) 975-3472	theodore.doiron@nist.gov	7

Service ID Number	Description of Services	Fee (\$)
11010S	Special Tests of Cylindrical Diameter Standards (i.e. Plug and Pin Gages): Set Up Charge, per order	150
11011S	Mechanical comparison, per Gage (25 mm and smaller)	91
11012S	Interferometry, per Gage (50 mm and smaller)	209
11013S	Per Gage (over 50 mm)	At Cost
11014S	Roundness trace, per trace	169
11020C	Measuring Wires for Threads and Gears: Set Up Charge, per order	157
11021C	Single Wire, per wire	95

11030S	Special Tests of Spherical Diameter Standards; Balls: Set Up Charge, per order (applies to mechanical comparison and interferometry)			
11031S	Mechanical Comparison (51 mm or smaller), Average Diameter, per ball, Expanded Uncertainty, U \sim \pm 89 nm to 115 nm	87		
11032S	Mechanical Comparison (over 51 mm), Average Diameter, per ball	At Cost		
11033S	Interferometry (25 mm or smaller) Average Diameter, per ball, Expanded Uncertainty, U $\sim\pm$ 30 nm	445		
11034S	Ball Out-of-Roundness: Least-Squares out-of-Roundness and Polar Plots, price per trace (Typically three orthogonal traces for spheres and five traces for CMM calibration spheres)			
	Special Tests of Internal Diameter Standards: Ring Gages			
11040S	Plain Ring Gages, per ring	720		
11050S	Special Tests of Length and Diameter	At Cost		
11060S	Special Tests of Step Gages	At Cost		

C. Complex Dimensional Standards

Technical Contacts:	Telephone:	Email:	Mailing Address:
Dennis Everett (12010C-12070S)	(301) 975-5272	dennis.everett@nist.gov	NIST
John Stoup (12060S)	(301) 975-3476	john.stoup@nist.gov	100 Bureau Drive, Stop 8211
Theodore Doiron	(301) 975-3472	theodore.doiron@nist.gov	Gaithersburg, MD 20899-8211

Service ID Number	Description of Services	
	API Threaded Plug and Ring Gages	
12010C	Spec 5, 1.005 inches to 7 5/8 inches	1609
12011C	Spec 5, 8 5/8 inches to 20 inches	2567
12012C	Buttress Casing, 4 1/2 inches to 9 5/8 inches	1933
12013C	Buttress Casing, 10 inches to 13 3/8 inches	2359
12014C	Buttress Casing, 16 inches to 20 inches	2527
12015C	Line Pipe, 1/8 inch to 6 inches (New)	1609
12016C	Line Pipe, 8 inches to 20 inches (New)	2131
12017C	Extreme Line Casing, 5 inches to 7 inches (New)	2710
12018C	Extreme Line Casing, 5 inches to 7 inches (Used)	1021
12019C	Extreme Line Casing, 7 7/8 inches to 10 inches (New)	3164
12021C	Extreme Line Casing, 7 7/8 inches to 10 inches (Used)	1299
12022C	Spec 7 (Rotary), NC 23-NC 61 (New)	2123

12023C	Spec 7 (Rotary), NC 70 (New)	2320
12024C	Spec 7 (Rotary), 2 3/8 inches to 4 1/2 inches, Reg. (New)	2130
12025C	Spec 7 (Rotary), 5 1/2 inches to 8 5/8 inches, Reg. (New)	2313
12026C	Spec 7 (Rotary), Any Type (Used)	945
12027C	Spec 11B (Sucker Rods) P1, P2 Pin Go P7, P8 Pin Go B1, B2 Box Go (NEW)	1109/per set
12028C	Spec 11B (Sucker Rods) P1, P2 Pin Go P7, P8 Pin Go B1, B2 Box Go (USED)	557/per set
12029C	Spec 11B (Sucker Rods) P3, P4 Pin Cone B3, B4 Box Cone (NEW)	1319/per set
12031C	Spec 11B (Sucker Rods) P3, P4 Pin Cone B3, B4 Box Cone (USED)	511/per set
12032C	Spec 11B (Sucker Rods) P5, P6 Pin Cone B5, B6 Box Cone (NEW)	900/per set
12033C	Spec 11B (Sucker Rods) P5, P6 Pin B5, B6 Box Cone (USED)	467/per set
12050S	Special Test of Threaded Plug and Ring Gages	At Cost
12060S	Special Tests of Two-Dimensional Gages	At Cost
12070S	Special Complex Dimensional Test by Prearrangement	At Cost

D. Optical Reference Planes and Roundness Standards

Technical Contacts: Telephone: Email: Mailing Address:

Eric S. Stanfield (301) 975-4882 eric.stanfield@nist.gov NIST

Theodore Doiron (301) 975-3472 theodore.doiron@nist.gov 100 Bureau Drive, Stop 8211 Gaithersburg, MD 20899-8211

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
13010S	Special Tests of Optical Reference Planes (Flats): Optical Flat, ≤152 mm (6"), Per Surface	1261
13011S	Special Tests of Optical Reference Planes (Flats): Optical Flat, 152 mm to 203 mm (8")	1643
13012S	Special Tests of Optical Reference Planes (Flats): Optical Flat, 203 mm to 304 mm	2159
13013S	Special Tests of Optical Reference Planes (Flats): Optical Flat, ≥ 304 mm (12")	2693
13014S	Special Tests of Optical Reference Planes (Flats): Three Flat Calibration	At Cost
13020S	Special Test of Roundness (Sphere and Hemisphere Types) Reversal Method: Radial Deviations form Best Fit Least-Squares Circle at 360 positions.	1067
13030S	Special Tests of Roundness Calibration Specimens	At Cost

Fees are subject to change without notice.

E. Angular Measurements

Technical Contacts: Telephone: Email: Mailing Address:

Bryon S. Faust (301) 975-4351 bryon.faust@nist.gov NIST

Theodore Doiron (301) 975-3472 theodore.doiron@nist.gov 100 Bureau Drive, Stop 8211

Gaithersburg, MD 20899-8211

Service ID Number	Description of Services	Fee (\$)
14010C	Angle Gage Blocks: Set Up Charge, per order	150
14011C	Angle Block, per block	258
14020S	Special Tests of Optical Polygons	At Cost
14030S	Special Tests of Rotary and Indexing Tables: Every 30°	2107
14031S	Special Tests of Rotary and Indexing Tables: (30°, 5°, 1°) Calibration	4110

14040S	Special Tests of Optical Wedges: Fixed-Angle Wedge	730
14041S	Special Tests of Optical Wedges: Variable-Angle Wedge	At Cost
14050S	Special Angular Measurements, by Prearrangement	At Cost

F. Laser Measurements

Technical Contact: Telephone: Email: Mailing Address:

Jack Stone (301) 975-5638 jack.stone@nist.gov NIST

100 Bureau Drive, Stop 8211 Gaithersburg, MD 20899-8211

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
14510S	Laser Frequency/Wavelength, Full Calibration	2323
14511S	Quick Check of Frequency/Wavelength at Laboratory Conditions	1240

Fees are subject to change without notice.

G. Surface Texture

Technical Contact:Telephone:Email:Mailing Address:C. Dewey Foreman, Jr.(301) 975-5836charles.foreman@nist.govNIST100 Bureau Drive, Stop 8211Gaithersburg, MD 20899-8211

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
15010C	Roughness Calibration Specimens	1342
15030C	Step Height Measurements	1342
15040S	Surface Roughness and Topography Special Tests	At Cost

CHAPTER 5

MECHANICAL MEASUREMENTS

A. Hydrometers

Technical Contacts:	Telephone:	Email:	Mailing Address:
----------------------------	-------------------	--------	------------------

Sherry Sheckels (301) 975-5940 sherry.sheckels@nist.gov NIST

John D. Wright (301) 975-5937 john.wright@nist.gov 100 Bureau Drive, Stop 8361 Gaithersburg, MD 20899-8361

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
16010C	Reference Standard Hydrometers	2000
16020S	Hydrometers Special Tests	At Cost

Fees are subject to change without notice.

B. Volume and Density

Technical Contacts:	<u>Telephone:</u>	Email:	Mailing Address:
Sherry Sheckels	(301) 975-5940	sherry.sheckels@nist.gov	NIST
John D. Wright	(301) 975-5937	john.wright@nist.gov	100 Bureau Drive, Stop 8361
			Gaithersburg, MD 20899-8361

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
17010C	Volume Standards	2000
17020C	Volume Standards > 380 L, 2 points	2999
17030C	Volume Standards > 380 L, 5 points	5999
17040S	Volume Special Tests	At Cost

C. Flow Measurements

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email</u> <u>Mailing Address:</u>

Gina Kline (301) 975-4813 gina.kline@nist.gov NIST

(Gas Flow and Hydrocarbon Flow) 100 Bureau Drive, Stop 8361

John D. Wright (301) 975-5937 john.wright@nist.gov Gaithersburg, MD 20899-8361

(Gas Flow and Water Flow)

Sherry Sheckels (301) 975-5940 sherry.sheckels@nist.gov

(Hydrocarbon Flow)

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
18010C	Gas Flow Meters	4500
18020C	Water Flow Meters	4499
18030C	Hydrocarbon Flow Meters	5498
18040C	Transfer Standards	At Cost
18050S	Gas Flow Special Tests	At Cost
18060S	Water Flow Special Tests	At Cost
18070S	Hydrocarbon Liquid Flow Special Tests	At Cost

Fees are subject to change without notice.

See 30063S Special Tests for Low-Gas-Flow Instrumentation

D. Flow Measurements at Cryogenic Temperatures

Technical Contact: Telephone: Email Mailing Address:

Michael Lewis (303) 497-3458 mlewis@boulder.nist.gov NIST

Fax: (303) 497-5224 325 Broadway, MC 838.09 Boulder, CO 80305-3328

Service ID Number	Description of Services	Fee (\$)
18800S	Special Tests of Cryogenic Liquid Flow	At Cost

E. Air Speed Measurements

Technical Contacts: Telephone: Email: Mailing Address:

J. Michael Hall (301) 975-5947 j.hall@nist.gov NIST

John D. Wright (301) 975-5937 john.wright@nist.gov 100 Bureau Drive, Stop 8361 Gaithersburg, MD 20899-8361

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
19010C	High Air Speed Instruments 1.3 m/s to 67 m/s (3 mph to 150 mph)	2500
19020C	Low Air Speed Instruments 0.3 m/s to 10.2 m/s (15 fpm to 2,000 fpm)	2500
19030S	High Air Speed special Tests	At Cost
19040S	Low Air Speed Special Tests	At Cost

F. Mass Standards

Technical Contacts:
Vincent LeeTelephone:
(301) 975-6453Email:
vincent.lee@nist.govMailing Address:
NIST

Zeina J. Jabbour (301) 975-4468 zeina.jabbour@nist.gov 100 Bureau Drive, Stop 8221

Gaithersburg, MD 20899-8221

Administrative and Logistics:

(301) 975-6624

Fax: (301) 417-0514

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
22010C	Weight Set (1 mg to 100 g)	11494
22020C	Weight Set (1 mg to 1 kg)	13297
22030C	Weight Set (2 kg to 30 kg)	5688
22040C	Single Weights (1 mg to 1 kg)	1398
22060C	Single Weights (2 kg to 30 kg)	1833
22080C	Single Weights (> 30 kg to 1200 kg, 2 double substitution weighings)	At Cost
22100C	Single Weights (> 1200 kg to 30,000 kg)	At Cost
22110C	Single Weights (> 30 kg to 1200 kg, calibrated in a weighing design)	At Cost
22130C	Single Weights for Dead Weight Pressure Testers 5.9 kg to 22.7 kg (13 lb to 50 lb)	968
22140C	Single Weights for Dead Weight Pressure Testers > 22.7 kg (> 50 lb)	At Cost
22150C	Single Weights for Dead Weight Pressure Testers < 5.9 kg (< 13 lb)	747
22170S	Special Mass Measurement Services	At Cost

G. Force Measurements

Technical Contacts:Telephone:Email:Mailing Address:Rick L. Seifarth(301) 975-6652ricky.seifarth@nist.govNIST

Thomas W. Bartel (301) 975-6461 tbartel@nist.gov 100 Bureau Drive, Stop 8222 Kevin L. Chesnutwood (301) 975-6653 kchesnutwood@nist.gov Gaithersburg, MD 20899-8222

Administrative and Logistics:

(301) 975-6624 **Fax:** (301) 417-0514

Service ID Number	Description of Services	Fee (\$)
23010C	Force Transducers to 112 540 N (25 300 lbf) 1 mode	
23020C	Extra observation	52
23030C	Additional bridges	747
23040C	Force Transducers to 112 540 N (25 300 lbf) 2 modes	4382
23050C	Extra observation	52
23060C	Additional bridges	674
23070C	Force Transducers 112 540 N to 498 201 N (25 300 lbf to 112 000 lbf) 1 mode	3070
23080C	Extra observation	52
23090C	Additional bridges	771
23100C	Force Transducers 112 540 N to 498 201 N (25 300 lbf to 112 000 lbf) 2 modes	5985
23110C	Extra observation	157
23120C	Additional bridges	1677
23130C	Force Transducers 498 205 N to 1 334 467 N (112 000 lbf to 300 000 lbf) 1 mode	6176
23140C	Extra observation	144
23150C	Additional bridges	1048
23160C	Force Transducers 498 205 N to 1 334 467 N (112 000 lbf to 300 000 lbf) 2 modes	9987
23170C	Extra observation	225
23180C	Additional bridges	1887
23190C	Force Transducers 1 334 471 N to 4 448 222 N (300 00 lbf to 1 000 000 lbf)	7481
23200C	Extra observation	157
23210C	Additional bridges	1349

23220C	Force Transducers 1 334 471 N to 4 448 222 N (300 00 lbf to 1 000 000 lbf) 2 modes	12063
23230C	Extra observation	210
23240C	Additional bridges	2409
23250C	Force Transducers over 4 448 222 N (1 000 000 lbf) compression only	At Cost
23260S	Special Tests of Force Transducers	At Cost

H. Vibration Measurements

Technical Contacts:	Telephone:	Email:	Mailing Address:
David J. Evans	(301) 975-6637	david.evans@nist.gov	NIST
Kari Harper	(301) 975-6612	kari.harper@nist.gov	100 Bureau Drive, Stop 8220
Toni Savoy	(301) 975-6613	toni.savoy@nist.gov	Gaithersburg, MD 20899-8220

Administrative and Logistics:

Annie Sokol (301) 975-3506 annie.sokol@nist.gov

Fax: (301) 975-8058

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Freq. Range	Peak Accel.	Fee (\$)
24010C	Transducer Sensitivity	2 Hz to 160 Hz	$0.2 g_n$ to $2 g_n$	2987
24020C	Transducer Sensitivity	10 Hz to 3500 Hz	2 g _n to 10 g _n	5180
24030C	Transducer Sensitivity	10 Hz to 10 kHz	2 g _n to 10 g _n	8495
24040S	Shock Measurement	250 Hz to 10 kHz	$20~g_n$ to $10~000~g_n$	At Cost
24050S	Transducer Sensitivity 3 kHz to 20 kHz 4 g _n to 200 g _n		At Cost	
24060S	S Special Vibration Tests, by Prearrangement			At Cost

Fees are subject to change without notice.

I. Acoustic Measurements

_	Technical Contacts:	Telephone:	Email:	Mailing Address:
•	Victor Nedzelnitsky	(301) 975-6638	vnedzelnitsky@nist.gov	NIST
I	Randall P. Wagner	(301) 975-6619	randall.wagner@nist.gov	100 Bureau Drive, Stop 8221
I	David J. Evans	(301) 975-6637	david.evans@nist.gov	Gaithersburg, MD 20899-8221

Service ID Number	Description of Services	Fee (\$)
25010C	Pressure Response: WE Type 640AA microphones or equivalent (e.g., Tokyo Riko Type ECL MR103; Bruel & Kjaer Type 4160, Bruel & Kjaer Types 4144 or 4132 with DB0111 adapter), 50 Hz to 10,000 Hz	5263

Service ID Number	Description of Services	Fee (\$)
25020C	Pressure Response: WE Type 640AA microphones or equivalent (e.g., Tokyo Riko Type ECL MR103; Bruel & Kjaer Type 4160; Bruel & Kjaer Types 4144 or 4132 with DB0111 adapter), 50 Hz to 20,000 Hz	6343
25030C	Pressure Response: Tokyo Riko Type ECL MR112, Bruel & Kjaer Type 4134, or equivalent half-inch microphones, 50 Hz to 10,000 Hz	6051
25040C	Pressure Response: Tokyo Riko Type EC MR112, Bruel & Kjaer Type 4134, or equivalent half-inch microphones, 50 Hz to 20,000 Hz	7603
25050C	Free-Field Response: Tokyo Riko Type ECL MR112, Bruel & Kjaer Types 4133, 4134, 4165, 4166, 4180, or equivalent half-inch microphones, 2,500 Hz to 20,000 Hz	6260
25060S	Special Test of Acoustic Devices	At Cost
25070S	Special Tests of Earphones	5381

CHAPTER 6 THERMODYNAMIC QUANTITIES

A. Pressure Measurements

Technical Contacts:Telephone:Email:Mailing Address:Douglas A. Olson (All Services)(301) 975-2956dolson@nist.govNIST

R. Gregory Driver (301) 975-4832 rdriver@nist.gov 100 Bureau Drive, Stop 8364 (pneumatic gages) Gaithersburg, MD 20899-8364

(29010C,29030C, 29035C, 29040S)

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
29010C	Deadweight Piston Gages	6404
29020C	Controlled Clearance Piston Gages	At Cost
29030C	Pressure Gages and Transducers	At Cost
29035C	Non-mercurial Barometers and Manometers	At Cost
29040S	Special Tests of Pressure Gages	At Cost

Fees are subject to change without notice.

B. Vacuum, Low Pressure and Leak Measurements

Technical Contacts:Telephone:Email:Mailing Address:Jay Hendricks (30010C-(301) 975-4836jay.hendricks@nist.govNIST

30025C, 30040S) 100 Bureau Drive, Stop 8364

Patrick J. Abbott (30029C- (301) 975-4838 patrick.abbott@nist.gov Gaithersburg, MD 20899-8364

30038C, 30050S, 30060S-30062C)

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

NOTE: 1 Torr = 133.322 Pa

Service ID Number	Description of Services	Fee (\$)
30010C	One Low-Pressure Transducer Absolute or Differential Relative to Vacuum	4242
30011C	Additional Transducer (Cost per Unit)	4057
30020C	One Differential Low-Pressure Transducer Relative to near Atmospheric Pressure	4732
30021C	Additional Transducers (Cost per Unit)	4265
30025C	Piston Gauges versus and Ultrasonic Interferometer Manometer	At Cost
30029C	Spinning Rotor Gages, below 0.1 Pa, Nitrogen Gas with NIST Controller	4064
30030C	Spinning Rotor Gages, below 0.1 Pa, Nitrogen Gas Customer Controller with IEEE-488	4064
30031C	Spinning Rotor Gages, below 0.1 Pa, Additional Gas	5025

30032S	Special Test of Spinning Rotor Gages, Transition Range (above 0.1 Pa)	At Cost
30034C	Ionization Gages, 10 ⁻⁴ Pa to 10 ⁻¹ Pa, Nitrogen Gas	4474
30035C	Ionization Gages, 10 ⁻⁵ Pa to 10 ⁻¹ Pa, Nitrogen Gas	5493
30036C	Ionization Gages, 10 ⁻⁷ Pa to 10 ⁻¹ Pa, Nitrogen Gas	6301
30037C	Ionization Gages, Additional Filament or Gas for Above Tests	At Cost
30038C	Ionization Gages, NIST Supplied Gage Tube for Above Tests	303
30040S	Special Tests of Low-Pressure Gages	At Cost
30050S	Special Tests of Vacuum Gages	At Cost
30060S	Special Tests of Leak Artifacts (10 ⁻¹³ mol/s to 10 ⁻⁶ mol/s)	At Cost
30061C	Helium Leaks, Primary Calibration (10 ⁻¹³ mol/s to 10 ⁻⁶ mol/s)	5548
30062C	Helium Leaks, Comparison Calibration (10 ⁻¹³ mol/s to 10 ⁻⁹ mol/s)	4382
30063S	Special Tests of Low-Gas-Flow Instruments	At Cost

NOTE: Due to the time and effort required preparing vacuum instrumentation for calibration it is particularly important that they be known to be in proper operating condition when they are submitted to NIST. Equipment will be inspected upon receipt and the customer notified of any obvious damage. If the schedule permits, we will cooperate with the customer's efforts to repair or replace damaged equipment so that the calibration of their equipment can proceed. However, concealed damage or operational deficiencies most likely will not be detected before the instrument is operating on the vacuum system or the calibration has started; in such cases, **if the equipment cannot be calibrated, we will charge 20% of the regular calibration fee for low-pressure transducers and 30% of the regular fee for spinning rotor and ionization gages.**

C. Laboratory and Industrial-Grade Thermometers

Technical Contact: C. Dawn Cross **Mailing Address: Telephone:** Email:

(301) 975-4822 dawn.cross@nist.gov **NIST**

100 Bureau Drive, Stop 8363 Gaithersburg, MD 20899-8363

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

NOTE: The minimum number of test points per thermometer is two. Fahrenheit ranges are not direct conversions of the Celsius ranges.

Service ID Number	Description of Services	Fee (\$)
31010C	Laboratory Thermometers (0 °C to 150 °C) (32 °F to 300 °F)	403/pt
31020C	Laboratory Thermometers (151 °C to 315 °C) (301 °F to 600 °F)	493/pt
31030C	Laboratory Thermometers (316 °C to 550 °C) (601 °F to 1022 °F)	493/pt
31040C	Laboratory Thermometers (-1 °C to -110 °C) (31 °F to -166 °F)	493/pt
31050C	Laboratory Thermometers (Liquid N ₂) (-196 °C or -321 °F)	403/pt
31100C	Quantity Tests of Liquid-In-Glass Thermometers	At Cost
31110S	Special Tests of Industrial Platinum Resistance Thermometers, Thermistor Thermometers, Digital Thermometers and Other Types of Thermometers (0 °C to 150 °C) (32 °F to 300 °F)	403/pt
31120S	Special Tests of Industrial Platinum Resistance Thermometers, Thermistor Thermometers, Digital Thermometers and Other Types of Thermometers (151 °C to 315 °C) (301 °F to 600 °F)	493/pt
31130S	Special Tests of Industrial Platinum Resistance Thermometers, Thermistor Thermometers, Digital Thermometers and Other Types of Thermometers (316 °C to 550 °C) (601 °F to 1022 °F)	493/pt
31140S	Special Tests of Industrial Platinum Resistance Thermometer, Thermistor Thermometers, Digital Thermometers and Other Types of Thermometers (-1 °C to -110 °C) (31 °F to -166 °F)	493/pt
31150S	Special Tests of Industrial Platinum Resistance Thermometers, Thermistor Thermometers, Digital Thermometers and Other Types of Thermometers (Liquid N ₂) (-196 °C or -321 °F)	403/pt
31170S	Special Tests of Calorimetric Thermometers	1571
31180S	Special Test of Beckmann Thermometers	1624
31190S	Additional copy of Table from Results of 31110S-31150S at a Later Date	354
31200S	Preliminary Examination of Ineligible Thermometer	79
31250S	Additional Copy of Report	79
31260S	Special Thermometry Services, by Prearrangement	At Cost

D. Thermocouples, Thermocouple Materials, Thermometer Indicators

Technical Contacts:Telephone:Email:Mailing Address:Christopher W. Meyer (32010C-32150S)(301) 975-4825cmeyer@nist.govNISTKaren Garrity(301) 975-4818kgarrity@nist.gov100 Bureau Drive, Stop 8363(32010C-32101C, 32150S)Gaithersburg, MD 20899-8363

C. Dawn Cross (301) 975-4822 dawn.cross@nist.gov (32110C-32147C)

СО	MPARED CALI	BRATIONS, TEMP	ERATURE MEASU	RED WITH THI	ERMOCOUPLE	
Service ID Number	TC Type	Temp Range °C	Points	Min. Length (mm)	Temp.	Fee (\$)
32010C	S	0 to 1450	1 °C or 1 °F Interv. Table	700	0 to 1100 1450	1137
32020C	R	0 to 1450	1 °C or 1 °F Interv. Table	700	0 to 1100 1450	1137
32030C	В	0 to 1750	1 °C or 1 °F Interv. Table	1000	0 to 800 800 to 1100 1450 1750	1709
32031C	В	800 to 1750	1 °C or 1 °F Interv. Table	1000	800 to 1100 1450 1750	1137
32040C	Е	0 to 1000	4 to 15	700	0 to 1000	1137
32041C	J	0 to 760	4 to 15	700	0 to 760	1137
32042C	K	0 to 1100	4 to 15	700	0 to 1100	1137
32043C	N	0 to 1100	4 to 15	700	0 to 1100	1137
32044C	T	0 to 400	4 to 15	700	0 to 400	1137
32050C	Comparison calibration, two point minimum, per point, for all items above					521/pt
32060C	Each additional table of results at 1 °C or 1 °F intervals, for type S, R, or B at later date					294
32061C	Each additional table of results at 1 °C or 1 °F intervals, for type S, R, or B at time of test					173
32070C	Thermocouple materials tested against Pt Thermoelectric standard, 4 to 15 points, 700 mm minimum lengths					1137

Service	TC	Temp	Points	Min.	Temp.	Fee (\$)
ID	Type	Range		Length	(°C)	
Number		$^{\circ}\mathrm{C}$		(mm)	. ,	
32090C	S or R	0 to 1450	Table 1 °C or 1 °F Interv. And equations to generate table	1000	at freezing points	2871
	K				0 to 1100 1450	
32091C	Type S or T	Γ, freezing point dete	ermination, per point, two poin	t minimum	•	1012
	•		CALIBRATION OF			
	DIGITAL	THERMOMETER	R INDICATOR OR PORTA	BLE POTENT	TOMETER	
32100C	Indicator or	Potentiometer, first	dial or range			857
32101C	Indicator or	Potentiometer, each	additional dial or range			477
PLAT	INUM RESIS'	TANCE THERMO	TANDARD, TEMPERATUR METER, MINIMUM TC W MINIMUM	TRE LENGTH	H 1.0 M, TWO PC	
32110C	Range -110 °C to 315 °C and Liquid N_2 (-196 °C) or -166 °F to 600 °F and Liquid N_2 (-321 °F), Expanded Uncertainty 0.4 °C					5'/'//pt
	(-321 °F), E	Expanded Uncertaint			iquiu 1 1/2	Зттре
32120C					iquid 112	577/pt
Fable at one	316 °C to 5	50 °C or 601 °F to 1	y 0.4 °C	v 0.5 °C	(The cost of the t	577/pt
Fable at one	316 °C to 5 degree intervon to the calibration 1: T	50 °C or 601 °F to 1 vals for Type T ther ration per point cov	y 0.4 °C 022 °F, Expanded Uncertainty rmocouple for any of the followered under fee schedule serv 0 + 300°C (-321 °F to +572 °F	v 0.5 °C owing options: vices numbered	(The cost of the t	577/pt
Fable at one	316 °C to 5 degree intervon to the calibration 1: T -110, 250, - Option 2: T	50 °C or 601 °F to 1 vals for Type T ther ration per point cov Table from -196 °C to +100, +200, +300) °C	y 0.4 °C 022 °F, Expanded Uncertainty rmocouple for any of the followered under fee schedule serv 0 + 300°C (-321 °F to +572 °F	owing options: vices numbered), calibration po	(The cost of the the discontinuous at (-196,	577/pt able will
Fable at one be in addition 32141C	Option 2: T -110, 250, -1 Option 3: T	750 °C or 601 °F to 1 Frais for Type T therefore the coverage of the coverage	y 0.4 °C 022 °F, Expanded Uncertainty rmocouple for any of the followered under fee schedule servents 0 + 300°C (-321 °F to +572 °F C	owing options: vices numbered), calibration po	(The cost of the the standard of the the standard of the stand	577/pt able will 3.
Table at one be in additional add	316 °C to 5 degree intervon to the calibration 1: T -110, 250, + Option 2: T -110, 250, + Option 3: T 250, +100,	750 °C or 601 °F to 1 Type T therefore the point coverage of the	y 0.4 °C 022 °F, Expanded Uncertainty rmocouple for any of the followered under fee schedule served 0 + 300°C (-321 °F to +572 °F C) 0 +100 °C (-321 °F to +212 °F)	owing options: vices numbered), calibration po	(The cost of the the standard of the the standard of the stand	577/pt able will 495
32141C 32142C 32143C	316 °C to 5 degree intervent to the calibration 1: T -110, 250, + Option 2: T -110, 250, + Option 3: T 250, +100, Option 4: T 250, +50, +	750 °C or 601 °F to 1 Type T therefore the point coverage of the	y 0.4 °C 022 °F, Expanded Uncertainty rmocouple for any of the followered under fee schedule server to +300°C (-321 °F to +572 °F) o +100 °C (-321 °F to +212 °F) o +300 °C (-166 °F to +572 °F)	outing options: vices numbered), calibration po), calibration po), calibration po), calibration po	(The cost of the total 32110C-32120C) oints at (-196, oints at (-196, oints at (-110, oints at	577/pt able will 495 495 495
32141C 32142C 32143C 32144C	316 °C to 5 degree intervent to the calibration of	zals for Type T ther ration per point coverage of the state of the sta	y 0.4 °C 022 °F, Expanded Uncertainty rmocouple for any of the followered under fee schedule served and the served of the solution of the followered under fee schedule served to +300°C (-321 °F to +572 °F C) o +100 °C (-166 °F to +572 °F C) o +100 °C (-166 °F to +212 °F C)	owing options: vices numbered), calibration po F), calibration po calibration po calibration po calibration po ration points at	(The cost of the total 32110C-32120C) oints at (-196, oints at (-196, oints at (-110, oints at (-110, (+100, +200,	577/pt able will 495 495
Table at one be in addition 32141C 32142C 32143C 32144C 32145C	316 °C to 5 degree intervent to the calibration to the calibration 1: T -110, 250, + Option 2: T -110, 250, + Option 3: T 250, +100, Option 4: T 250, +50, + Option 5: T +300) °C Option 6: T	zals for Type T ther ration per point coverage in the state of the sta	y 0.4 °C 022 °F, Expanded Uncertainty mocouple for any of the followered under fee schedule served to +300°C (-321 °F to +572 °F) o +100 °C (-321 °F to +212 °F) o +100 °C (-166 °F to +572 °F) o +100 °C (32 °F to +572 °F), calib	owing options: vices numbered), calibration po F), calibration po Calibration po Calibration po Calibration po Calibration po Calibration po Calibration points at	(The cost of the total 32110C-32120C) oints at (-196, oints at (-196, oints at (-110, oints at (-110, (+100, +200, at (-110, 250) °C	577/pt able will 2). 495 495 495 495

NOTE: Due to the extra time involved in calibrating sheathed thermocouples, a surcharge of 20% of the cost of calibrating bare-wire thermocouples will be added to the relevant fees listed above.

E. Resistance Thermometry

Technical Contacts:
Weston L. Tew (0.65 K to 84 K)

Gregory F. Strouse (83K to 962 °C)

<u>Telephone:</u> (301) 975-4811 (301) 975-4803

Email: wtew@nist.gov gstrouse@nist.gov Mailing Address: NIST

100 Bureau Drive, Stop 8363 Gaithersburg, MD 20899-8363

Service ID Number	Description of Services	Fee (\$)
33010C	Capsule SPRT (13.8 K to 30 °C) e-H ₂ to Ga	9917
33020C	Capsule SPRT (13.8 K to 157 °C) e-H ₂ to In	9917
33030C	Capsule SPRT (13.8 K to 232 °C) e-H ₂ to Sn	10893
33031C	Capsule SPRT (24.5 K to 30 °C) Ne to Ga	7799
33032C	Capsule SPRT (24.5 K to 157 °C) Ne to Ga	8224
33033C	Capsule SPRT (24.5 K to 232 °C) Ne to Sn	9200
33040C	Capsule SPRT(54 K to 30 °C) 0 ₂ to Ga	6969
33050C	Capsule SPRT (54 K to 157 °C) 0 ₂ to In	7395
33060S	Capsule SPRT (54 K to 232 °C) 0 ₂ to Sn	9093
30065S	Capsule SPRT (83 K to 0.01 °C) Ar to TPW	At Cost
30070C	Capsule SPRT (83 K to 30 °C) Ar to Ga	6896
33080C	Capsule SPRT (83 K to 157 °C) Ar to In	6896
33090C	Capsule SPRT (83 K to 232 °C) Ar to Sn	8767
33100C	Capsule SPRT (0 °C to 30 °C) TPW to Ga	2624
33110C	Capsule SPRT (0 °C to 157 °C) TPW to In	3049
33120C	Capsule SPRT (0 °C to 232 °C) TPW to Sn	4748
33130C	Capsule SPRT (234 K to 30 °C) Hg to Ga	4824
33140C	Rhodium-Iron or Platinum-Cobalt Resistance Thermometers (0.65 K to 24.6 K)	10424
33141C	Rhodium-Iron or Platinum-Cobalt Resistance Thermometers (0.65 K to 83.8 K)	12829
33142C	n-Type Germanium Resistance Thermometers (0.65 K to 24.6 K)	11781
33150C	Long Stem SPRT (83 K to 0.01 °C) Ar to TPW	3499
33160C	Long Stem SPRT (83 K to 30 °C) Ar to Ga	4136
33170C	Long Stem SPRT (83 K to 157 °C) Ar to In	4815
33180C	Long Stem SPRT (83 K to 232 °C) Ar to Sn	5410
33190C	Long Stem SPRT (83 K to 420 °C) Ar to Zn	6471

33200C	Long Stem SPRT (83 K to 661 °C) Ar to Al	7576
33210C	Long Stem SPRT (234 K to 30 °C) Hg to Ga	2871
33220C	Long Stem SPRT (234 K to 157 °C) Hg to In	3763
33230C	Long Stem SPRT (234 K to 232 °C) Hg to Sn	4621
33240C	Long Stem SPRT (234 K to 420 °C) Hg to Zn	5590
33250C	Long Stem SPRT (234 K to 661 °C) Hg to Al	6650
33260C	Long Stem SPRT (0 °C to 30 °C) TPW to Ga	1400
33270C	Long Stem SPRT (0 °C to 157 °C) TPW to In	2020
33280C	Long Stem SPRT (0 °C to 232 °C) TPW to Sn	2665
33290C	Long Stem SPRT (0 °C to 420 °C) TPW to Zn	3515
33300C	Long Stem SPRT (0 °C to 661 °C) TPW to Al	4748
33310C	Long Stem SPRT (0 °C to 962 °C) TPW to Ag	9595
33320C	Additional Copy of Table from Results of 33010C-33310C at Time of Test	117
33330C	Additional Copy of Table from Results of 33010C-33310C at a Later Date	347
33340C	Minimum Charge for Unsuitable Thermometer	324
33350S	Special Tests of Resistance Thermometers	At Cost
33355S	Special Tests of Cryogenic Resistance Thermometers	At Cost
33360S	Special Tests of Thermometric Fixed-Point Devices	At Cost
33370M	Measurement Assurance Program for Temperature 83 K to 420 °C (Ar to Zn)	22516
33380M	Measurement Assurance Program for Temperature 83 K to 661 °C (Ar to Al)	24786

F. Radiance Temperature Measurements

Technical Contact: Charles E. Gibson Telephone:Email:(301) 975-2329cgibson@nist.gov **Mailing Address:**

NIST

100 Bureau Drive, Stop 8441 Gaithersburg, MD 20899-8441

Fax: (301) 869-5700

Service ID Number	Description of Services	Fee (\$)
Calibration	reports are issued giving the radiance temperature of the blackbody at 655.48 the scale reading, output current, or output voltage	nm versus
35010C	Radiance Temperature Standard, Disappearing Filament Optical Pyrometer (800 °C to 2400 °C, 4 to 12 points, 1 range)	7093
35020C	Radiance Temperature Standard, Disappearing Filament Optical Pyrometers (each additional range up to 4200 °C, only available with 35010C)	5123
35040C	Radiance Temperature Standard, Disappearing Filament Optical Pyrometer (800 °C to 4200 °C, 1 range 3 or fewer points)	3547
Calibration	n reports are issued giving the radiance temperature of the lamp at 655.48 nm slamp current	versus the
35050C	Radiance Temperature Standard, Tungsten Strip Lamp (800 °C to 2300 °C, 6 to 16 points)	9853
35051C	Recalibration of Tungsten Strip Lamp (800 °C to 2300 °C, 6 to 16 points)	8277
35060C	Radiance Temperature Standard, Tungsten Strip Lamp (800 °C to 2300 °C, 5 or fewer points)	6306
35061C	Recalibration of Tungsten Strip Lamp (800 °C to 2300 °C, 5 or fewer points)	4729
	tion reports are issued giving the radiance temperature of the reference blackb nm, 900 nm or 1000 nm versus the display reading, output current, or output v	
35070S	Special Tests of Radiation Thermometers (800 °C to 2700 °C)	At Cost
35071C	Radiance Temperature Standard, Radiation Thermometer (800 °C to 2700 °C, 6 to 20 points)	7488
35072C	Radiance Temperature Standard, Radiation Thermometer (800 °C to 2700 °C, 5 or fewer points)	3941
Calibratio	n reports are issued giving the thermodynamic temperature of the reference by versus the display reading, output current, or output voltage.	lackbody
35080S	Special Tests of Radiation Thermometers (15 °C to 900 °C)	At Cost
35081C	Radiance Temperature Standard, Radiation Thermometer (15 °C to 70 °C, 3 points)	3941
35082C	Radiance Temperature Standard, Radiation Thermometer (70 °C to 170 °C, 3 points)	3941

35083C	Radiance Temperature Standard, Radiation Thermometer (400 °C to 700 °C, 3 points)	3941			
35084C	Radiance Temperature Standard, Radiation Thermometer (700 °C to 900 °C, 3 points)	3941			
Calibratio	Calibration reports are issued giving the thermodynamic temperature of the reference blackbody versus the test blackbody source display reading.				
35090S	Special Tests of Blackbody Sources (15 °C to 900 °C)				
Calibrat	Calibration reports are issued giving heat flux at the sensor surface versus the output voltage.				
35100S	Special Tests of Radiative Heat Flux Sensors	At Cost			
35101C	Radiative Heat Flux Sensors (1 W/cm ² to 5 W/cm ² , 9 points, Gardon and Schmidt-Boelter type sensors)	3153			
35102C	Additional Radiative Heat Flux Sensor (same model as 35101C)	2365			

Calibration Schedule: Requests for calibration services are scheduled after receipt of a purchase order.

G. Humidity Measurements

Technical Contacts:	Telephone:	Email:	Mailing Address:
Peter H. Huang	(301) 975-2621 or 2626	phuang@nist.gov	NIST
Joseph T. Hodges	(301) 975-2605	jhodges@nist.gov	100 Bureau Drive, Stop 8363
Gregory E. Scace	(301) 975-2626	gregory.scace@nist.gov	Gaithersburg, MD 20899-8363
			Fax: (301) 548-0206

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
36010C	Dew-Point Hygrometers (+25 °C to -15 °C)	5907
36020C	Dew-Point Hygrometers (-70 °C to -15 °C)	11199
36030C	Electric Hygrometers	At Cost
36040C	Electrolytic Hygrometers	At Cost
36050C	Aspirated Hygrometers	At Cost
36060C	Pneumatic Bridge Hygrometers	At Cost
36070S	Special Tests of Humidity	At Cost

CHAPTER 7

OPTICAL RADIATION MEASUREMENTS

A. Photometric Measurements

Technical Contact: Telephone: Email: Mailing Address:

Cameron Miller (301) 975-4713 c.miller@nist.gov NIST

100 Bureau Drive, Stop 8442 Gaithersburg, MD 20899-8442

Fax: (301) 840-8551

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services		
37010C	Luminous Intensity and Color Temperature Standard Lamps	4212	
37020S	Special Tests for luminous Intensity and Color Temperature of Submitted Lamps	At Cost	
37030C	Color Temperature Standard Lamps	3423	
37040C	Each Additional Color Temperature for 37030C	788	
37050S	Special Tests for Color Temperature of Submitted Lamps	At Cost	
37060S	Special Tests for Total Luminous Flux of Submitted Incandescent Lamps and Florescent Lamps	At Cost	
37070C	Opal Glass Luminance Coefficient Standards	3028	
37080S	Special Tests for Submitted Luminance Sources and Transmitting Diffusers	At Cost	
37090S	Special Tests for Photometers, Illuminance Meters and Luminance Meter	At Cost	
37100S	Special Photometric Tests	At Cost	
37110S	Special Tests for Submitted Flashing-Light Photometers	At Cost	
37120S	Special Tests for Color Measuring Instruments for Displays	At Cost	
37130S	Special Tests for Luminous Intensity and Luminous Flux of LEDs	At Cost	

B. Optical Properties of Materials Measurements

(38010C-38091S)

Technical Contacts:	Telephone:	Email:	Mailing Address:
David W. Allen	(301) 975-3680	david.allen@nist.gov	NIST
Maria E. Nadal	(301) 975-4632	maria.nadal@nist.gov	100 Bureau Drive, Stop 8442

Gaithersburg, MD 20899-8442

Fax: (301) 840-8551

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
38010C	Spectral Transmittance Filters (Cobalt Blue Glass)	3571
38020C	Spectral Transmittance Filters (Copper Green Glass)	3571
38030C	Spectral Transmittance Filters (Carbon Yellow Glass)	3571
38040C	Spectral Transmittance Filters (Selenium Orange Glass)	3571
38050C	Wavelength Standards (Holmium Oxide Glass)	2945
38051C	Wavelength Standards	2945
38060S	Special Tests of Spectral Reflectance	At Cost
38061S	Special Tests of Spectral Transmittance	At Cost
38090S	Specular Gloss	At Cost
38091S	Special Test of 0°/45° Surface Color	At Cost

Fees are subject to change without notice.

C. Spectroradiometric Measurements

Technical Contacts:	Telephone:	Email:	Mailing/Shipping Address:
Charles E. Gibson	(301) 975-2329	cgibson@nist.gov	NIST
(39010C-39060S)			100 Bureau Drive, Stop 8441
Jeanne M. Houston	(301) 975-2327	jeanne.houston@nist.gov	Gaithersburg, MD 20899-8441
(39071C-39081S)			Fax: (301) 869-5700
Thomas C. Larason	(301) 975-2334	tlarason@nist.gov	
(39080S, 39081S, 39100S)			
George Eppeldauer (39090S)	(301) 975-2338	geppeldauer@nist.gov	

C.1 Spectroradiometric Source Measurements				
Service ID Number	Description of Services	Fee (\$)		
NIST calibrates and issues a type 30A/T24/13 tungsten strip lamp with a mogul bi-post base.				
39010C	Spectral Radiance Standard, Tungsten Strip Lamp (225 nm to 2400 nm) (other spectral ranges are available under no. 39060S)	13981		

11101 0	alibrates customer supplied integrating sphere sources and maps the source ape	erture.
39020C	Spectral Radiance Standard, Integrating Sphere Source (300 nm to 1000 nm in 25 nm steps)	7698
39021C	Spectral Radiance Standard, Integrating Sphere Source (300 nm to 2400 nm in 25 nm steps)	10851
NIST o	calibrates and issues an 1000 W, tungsten quartz-halogen lamp mounted in a me bi-post base. The calibrations are performed at 50 cm.	edium
39030C	Spectral Irradiance Standard, 1000 W Tungsten Quartz-Halogen Lamp (250 nm to 450 nm)	10203
39031C	Recalibration of 1000 W Tungsten Quartz-Halogen Lamp (250 nm to 450 nm)	7119
39032C	Spectral Irradiance Standard, 1000 W Tungsten Quartz-Halogen Lamp (350 nm to 800 nm)	10203
39033C	Recalibration of 1000 W Tungsten Quartz-Halogen Lamp (350 nm to 800 nm)	7119
39040C	Spectral Irradiance Standard, 1000 W Tungsten Quartz-Halogen Lamp (250 nm to 1600 nm)	1277
39041C	Recalibration of 1000 W Tungsten Quartz-Halogen Lamp (250 nm to 1600 nm)	9694
39045C	Spectral Irradiance Standard, 1000 W Tungsten Quartz-Halogen Lamp (250 nm to 2400 nm)	15059
39046C	Recalibration of 1000 W Tungsten Quartz-Halogen Lamp (250 nm to 2400 nm)	12059
NIST	calibrates and issues a 30 W deuterium arc lamp mounted in a medium bi-post	base.
39050C	Spectral Irradiance Standard, 30W Deuterium Arc Lamp (200 nm to 400 nm)	13033
39051C	Recalibration of 30 W Deuterium Arc Lamp (200 nm to 400 nm)	9484
39060S	Special Tests of Radiometric Sources	At Cos
	C.2 Spectroradiometric Detector Measurements	
39071C	UV Silicon Photodiodes	4402
39072C	Recalibration of UV Silicon Photodiodes	3355
39073C	Visible to NIR Silicon Photodiodes	4494
39074C	Recalibration of Visible to NIR Silicon Photodiodes	3355
39075S	Special Tests of NIR Photodiodes	At Cos
39077C	UV to Near-Infrared Silicon Photodiodes (Hamamatsu S2281)	533
39078C	Recalibration of UV to Near-Infrared Silicon Photodiodes (Hamamatsu S1337-1010BQ or S2281)	4193
39080S	Special Tests of Radiometric Detectors	At Cos
39081S	Special Tests of Photodetector Responsivity Spatial Uniformity	At Cos
39090S	Special Tests of IR Detectors	At Cos
	1	

D. Radiometric Standards in the Ultraviolet

Standard Detectors in the Far Ultraviolet

<u>Technical Contact:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Robert E. Vest (301) 975-3992 rvest@nist.gov NIST

100 Bureau Drive, Stop 8411 Gaithersburg, MD 20899-8411

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
40510C	Detector Standard, Windowless Photodiode (5 nm to 122 nm)	3727
40511C	Recalibration of Detector Standard (5 nm to 122 nm)	3358
40520C	Detector Standard, Windowless Photodiode (18 nm to 122 nm)	2645
40521C	Recalibration of Detector Standard (18 nm to 122 nm)	2277
40530C	Detector Standard, Windowless Photodiode (52 nm to 122 nm)	1564
40531C	Recalibration of Detector Standard (52 nm to 122 nm)	1195
40540C	Uncalibrated Windowless Photodiode	683
40560C	Detector Standard, Windowless Photodiode (116 nm to 254 nm)	12214
40561C	Recalibration of Detector Standard (116 nm to 254 nm)	1195
40599S	Special Tests on Detectors from the Ultraviolet (254 nm) to the Soft X-Ray Region (5 nm)	At Cost

E. Laser and Optoelectronic Components Used with Lasers

Technical Contacts:	Telephone:	Email:	Mailing Address:	
John H. Lehman	(303) 497-3654	lehman@boulder.nist.gov	NIST	
(CW Laser Radiometry)			325 Broadway, MC 815.01	
Paul D. Hale	(303) 497-5367	hale@boulder.nist.gov	Boulder, CO 80305-3328	
(High Speed Measurements)				
Christopher L. Cromer	(303) 497-5620	cromer@boulder.nist.gov		
(Pulsed-Laser Radiometry)				
Timothy Drapela	(303) 497-5858	drapela@boulder.nist.gov		
(Optical Fiber and Component Measurements [other than Fiber Power])				

Administrative and Logistics:

John Lomax (303) 497-4285 john.lomax@boulder.nist.gov

FAX: (303) 497-4286

Service ID Number	Description of Services	Fee (\$)
42110C	Laser Power and Energy Meter (or Detector) Calibrations at a Single Standard W and Power (See Table 4)	avelength
	CW Laser Power below 2 Watts	3476
	Pulsed Laser Energy (Q-switched YAG) at 1064 nm	3596
	CW Laser Power at 1064 nm above 2 Watts and 10.6 μm	4494
	Pulsed Laser Energy (Excimer) at 248 nm and 193 nm	4608
42111C	Same as 42110C, Additional Standard Wavelengths or Powers (See Table 4)	
	CW Laser Power below 2 Watts	1738
	Pulsed Laser Energy (Q-switched YAG) at 1064 nm	2697
	CW Laser Power at 1064 nm above 2 Watts and 10.6 µm above 1 Watt	3596
	Pulsed Laser Energy (Excimer) at 248 nm and 193 nm	3411
42120M	Laser Power and Energy Measurement Assurance Program (MAP)	At Cost
42130C	Optical Fiber Power Meter (or Detectors Used with Lasers) Calibrations at a Single Standard Wavelength and Connector Type (See Table 5)	2504
42131C	Same as 42130C, Additional Standard Wavelengths or Connector Types (See Table 5)	1002
42140M	Optical Fiber Power Meter Measurement Assurance Program (MAP)	At Cost
42150M	Low-Level Laser Measurement Assurance Program (MAP)	At Cost
42151C	Low-Level Laser Radiometer Calibration	At Cost
42155C	Calibration Service of Optoelectronic Frequency Response for Combined Photodiode/RF Power Sensor Transfer Standards	At Cost
42160S	Special Test for Frequency Response Measurements of Detectors Used with Lasers	At Cost

42161S	Special Test for Impulse Response Measurements of Detectors Used with Lasers	At Cost
42162S	42162S Special Test for High Accuracy Laser and Optical Fiber Power Measurements	
42164C	Spectral Responsivity Measurements of Laser and Optical Fiber Power Meters (or Detectors Used with Lasers)	3107
42165S	Special Test for Spatial Uniformity of Laser and Optical Fiber Power Meters and Detectors Used with Lasers	At Cost
42166C	42166C Calibration for Linearity Measurements of Optical Fiber Power Meters (or Detectors Used with Lasers) 42167S Special Test for Linearity Measurements of High-Power Laser Power Meters (or Detectors Used with Lasers) 42170S Special Test for General Laser Measurements, by Prearrangement 42180S Special Test for General Optical Fiber Power Measurements, by Prearrangement 42190S Special Test for Optical Fiber and Fiber Component Measurements (other than Fiber Power) by Prearrangement	
42167S		
42170S		
42180S		
42190S		
42210C Spectral Responsivity Measurements with Curve Fitting of Laser and Optical Meters (or Detectors used with Lasers)		3455

CHAPTER 8

IONIZING RADIATION MEASUREMENTS

A. Radioactivity Sources

Technical Contacts:	Telephone:	Email:	Mailing Address:
Lisa R. Karam	(301) 975-5561	lisa.karam@nist.gov	NIST
(All Services)			100 Bureau Drive, Stop 8462
M.P. Unterweger	(301) 975-5536	munterweger@nist.gov	Gaithersburg, MD 20899-8462
(43030C,43040C,			Attn: Jeffrey Cessna
43070S, 43090S)			
Jeffrey T. Cessna	(301) 975-5539	jcessna@nist.gov	
(43010C, 43020C,			
43060C, 43070S)	(201) 075 5544		
Lynne King	(301) 975-5544	lynne.king@nist.gov	
(43030C, 43040C,			
43070S, 43090S)			

Administrative and Logistics: Jeffrey Cessna (301) (301) 975-5539 jcessna@nist.gov

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
43010C	Gamma-Ray-Emitting Radionuclides in Solution (Half Lives Greater than 15 Days)	2307
43020C	Gamma-Ray-Emitting Radionulcides in Solution (Half Lives Less than 15 Days)	3862
43030C	Alpha-Particle-Emitting Solid Sources, NIST 2 π∞ Proportional Counter	1611
43040C Alpha-Particle-Emitting Solid Sources, NIST 0.8 π∞ Defined-Solid-Angle Counter		1595
43050C	Alpha-Particle-Emitting Solid Sources, Using Both Counting Systems	
43060S	Special Tests of Beta-Particle-Emitting Solution Sources, Liquid Scintillation Counting	4726
43070S	Special Tests of Beta-Particle-Emitting Solution Sources, Other Techniques	At Cost
43090S	Special Tests of Alpha-Particle-Emitting Solid Sources	At Cost

B. Neutron Sources and Neutron Dosimetry

<u>Technical Contacts:</u>	Telephone:	Email:	Mailing Address:
M. Scott Dewey	(301) 975-4843	mdewey@nist.gov	NIST
(All Services Except 44060C)			100 Bureau Drive, Stop 8461
Alan K. Thompson	(301) 975-4666	alan.thompson@nist.gov	Gaithersburg, MD 20899-8461
(44060C, 44100S)			

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	
44010C	Radioactive Neutron Sources Emission Rates (10 ⁵ s ⁻¹ to 10 ⁹ s ⁻¹)	5113
44020C	Radioactive Neutron Sources Emission Rates (10 ⁸ s ⁻¹ to 10 ¹⁰ s ⁻¹)	5113
44060C	Personnel Protection Instrumentation, Californium Source Bare and Moderated	
44070C	Activation Detector Dosimetry, Thermal Neutrons	
44080C	Activation Detector Dosimetry, Californium Fission Neutrons	At Cost
44090C	Activation Detector Dosimetry, ²³⁵ U Cavity Fission Sources	At Cost
44100S	Special Test of Neutron Sources and Dosimeters	At Cost

Fees are subject to change without notice.

C. Dosimetry of X-Rays, Gamma-Rays, and Electrons

Technical Contacts:	Telephone:	Email:	Mailing Address:
Stephen M. Seltzer (All Services)	(301) 975-5552	s.seltzer@nist.gov	NIST
Michael Mitch	(301) 975-5491	mmitch@nist.gov	100 Bureau Drive, Stop 8460
(46010C-47040S)			Gaithersburg, MD 20899-8460
Michelle O'Brien	(301) 975-2014	michelle.obrien@nist.gov	Fax: (301) 869-7682
(46010C-46050S)			
Ronaldo Minniti	(301) 975-5586	ronaldo.minniti@nist.gov	
(46010C-46110C)			
Heather Chen-Mayer	(301) 975-5595	heather.chen-mayer@nist	.gov
(46110C)			
Christopher G. Soares	(301) 975-5589	csoares@nist.gov	
(47030C-47040S)			

	C.1 X-Ray and Gamma-Ray Measuring Instruments		
Service ID Number	r		
	Air-Kerma (Exposure)		
46010C	Radiation Detectors—Calibration in ⁶⁰ Co and ¹³⁷ Cs Gammy-Ray Beams	1721	
46011C	46011C Radiation Detectors—Calibration in X-Ray Beams (see Tables 6, 7 and 8)		
46020C	Passive Dosimeters—Irradiation of Up to Six, One Beam Quality at One Set-up	1713	
46021C	Up to Six Additional Dosimeters at Same Set-up and Beam Quality	1083	

46030S	Special Tests of High-Gain Electrometers—Charge Sensitivity, One Set of Switch Positions, with 46010C/46011C, by Prearrangement	1131
46040S	Special Tests of kV Measuring Devices	At Cost
46050S	Special Tests of X-Ray and Gamma-Ray Measuring Instruments	At Cost
	Absorbed Dose to Water From ⁶⁰ Co Beam	
46110C	Radiation Detectors—Calibration in a 60Co Gamma-Ray Beam	2009
C.2 S	Sealed Gamma-Ray Sources or Beta-Particle Sources, and Measuring Instrume	ents
47010C	Gamma-Ray Sources Similar to NIST Standards— 60 Co to 137 Cs, Having Air-Kerma Strengths 10 μ Gy m²/h to 1500 μ Gy m²/h; and 192 Ir Sources of the Same Type Used to Calibrate Reentrant Chamber, Having Air-Kerma Strengths 0.1 μ Gy m²/h to 30 μ Gy m²/h	3220
47011C	Each Additional Gamma-Ray Source of Same Radionuclide	3111
47020C	125 I or 103 Pd Sources: Seeds Having Air-Kerma Strengths 0.5 $\mu \rm{Gy~m^2/h}$ to 100 $\mu \rm{Gy~m^2/h}$	3204
47021C	Each Additional ¹²⁵ I or ¹⁰³ Pd Source of Same Radionuclide/Design Submitted with Above	3111
47030C	Beta-Particle Source Calibrated for Surface Dose Rate	2105
47035C	Beta-Particle Source Calibrated for Radiation Protection	1679
47036C	Ionization Chamber Calibrated with Beta-Particle Sources for Radiation Protection	1679
47040S	Special Tests of Gamma-Ray and Beta-Particle Sources	At Cost

D. Dosimetry for High-Dose Applications

Technical Contacts:	Telephone:	Email:	Mailing Address:
Stephen M. Seltzer	(301) 975-5552	s.seltzer@nist.gov	NIST
Marc D. Desrosiers	(301) 975-5639	mdesrosiers@nist.gov	100 Bureau Drive, Stop 8460
James M. Puhl	(301) 975-5581	jpuhl@nist.gov	Gaithersburg, MD 20899-8460
Sarenee L. Cooper	(301) 975-5054	sarenee.cooper@nist.gov	

Service ID Number	Description of Services	Fee (\$)
	D.1 Dosimetry of High-Energy Electron Beams	
48010M	48010M Dose Interpretation of NIST-Packaged Dosimeters Irradiated by Customer— Two Dosimeters	
48011M	Each Additional Dosimeter	472
48020S	Special Tests of Electron-Beam Dosimeters	At Cost

	D.2 Dosimetry of Photon Beams		
49010C	49010C Calibration Irradiations of Customer Supplied Dosimeters with ⁶⁰ Co Gamma-Rays 49020C Dose Interpretation of NIST Transfer Dosimeters Irradiated by Customer, Three Dosimeters Plus Control(s)		
49020C			
49030C	Dose Interpretation of Each NIST Transfer Dosimeter Package in Addition to Those Supplied Under 49020C	625	
49050S	49050S Special Measurement Services for Dosimeter Response and Dose Distributions		

CHAPTER 9

ELECTROMAGNETIC MEASUREMENTS

A. Resistance Measurements

A.1 DC Resistance Standards and Measurements

Technical Contacts: Telephone: Email: Mailing Address:

George R. Jones (301) 975-4225 george.jones@nist.gov NIST

Randolph E. Elmquist (301) 975-6591 relmquist@nist.gov 100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
51100S	Special Resistance Measurements Services, by Prearrangement	At Cost
51110M	Measurement Assurance Program for Resistance	At Cost
51130C	Standard Resistor, Thomas-Type, 1 Ω	1777
51131C	Standard Resistor, Evanohm Wirewound High Precision, $10 \text{ k}\Omega$	1701
51132C	Standard Resistor, Four-Terminal 0.0001 Ω	1923
51133C	Standard Resistor, Four-Terminal 0.001 Ω	1652
51134C	Standard Resistor, Four-Terminal 0.01 Ω	1652
51135C	Standard Resistor, Four-Terminal 0.1 Ω	1247
51136C	Standard Resistor, Four-Terminal 1 Ω	1247
51137C	Standard Resistor, Four-Terminal 10 Ω	1247
51138C	Standard Resistor, Four-Terminal $100~\Omega$	1247
51139C	Standard Resistor, 1 kΩ	1247
51140C	Standard Resistor, $10 \text{ k}\Omega$	1605
51141C	Standard Resistor, 100 kΩ	1605
51142C	Standard Resistor, 1 $M\Omega$	1784
51143C	Standard Resistor, $10 \text{ M}\Omega$	1959
51144C	Additional Voltage, $10~\text{M}\Omega$	1605
51145C	Standard Resistor, $100 \text{ M}\Omega$	1959
51146C	Additional Voltage, $100 \text{ M}\Omega$	1605
51147C	Standard Resistor, 1 G Ω	1959
51148C	Additional Voltage, 1 GΩ	1605

51149C	Standard Resistor, $10 \text{ G}\Omega$	2497
51150C	Additional Voltage, $10~\mathrm{G}\Omega$	2145
51151C	Standard Resistor, $100~\text{G}\Omega$	2497
51152C	Additional Voltage, 100 GΩ	2145
51153C	Standard Resistor, 1 $T\Omega$	2632
51154C	Additional Voltage, 1 TΩ	2550
51160C	Standard Resistor for Current Measurements (Shunts) with all determinations at 300 A or Below, One Range, One Current Level	2771
51161C	Standard Resistor for Current Measurements (Shunts), with At Least One Determination Above 300 A (maximum current 2000 A), One Range, One Current Level	3403
51162C	Standard Resistor for Current Measurements (Shunts), Additional Range of a Multi-Range Resistor	1820
51163C	Standard Resistor for Current Measurements (Shunts), Additional Determination at Another Current Level	1820

A.2 High-Voltage Standard Resistors

Technical Contacts:	Telephone:	Email:	Mailing Address:
Gerald J. FitzPatrick	(301) 975-8922	gfitzpatrick@nist.gov	NIST
			100 Bureau Drive, Stop 8170

Gaithersburg, MD 20899-8170

Administrative and Logistics: Denise D. Prather (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
51210C	High-Voltage Standard Resistors	At Cost

A.3 High-Frequency Standard Resistors

Technical Contact: Telephone: Email: Mailing Address:

Ronald A. Ginley (303) 497-3634 rginley@boulder.nist.gov NIST

325 Broadway, MC 818.01 Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
51310S	High-Frequency Standard Resistors; Two-Terminal	At Cost

B. Impedance Measurements (Except Resistors)

B.1 Low-Frequency Capacitance and Inductance Measurements and Standards

Technical Contacts: Telephone: Email: Mailing Address:

Andrew D. Koffman (301) 975-4518 akoffman@nist.gov NIST

100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
52100S	Special Four Terminal-Pair (4TP) Capacitance and Dissipation Factor Characterization	At Cost
52110S	Special LF Capacitance Measurements, by Prearrangements	At Cost
52120S	Special Measurement Assurance Program for Standard Capacitors (100 pF and 1000 pF, at a Frequency of 1000 Hz)	At Cost
52130C	Fixed, Fused-Silica Dielectric Standard Capacitors (1, 10, and 100) pF, at a Frequency of (100, 400, or 1000) Hz	2660
52131C	Additional Measurement at One of the Above Frequencies	278
52140C	Fixed Three-Terminal, High-Precision Nitrogen Dielectric Standard Capacitors with Coaxial Connectors, Small Uncertainty, (10, 100 and 1000) pF, at a Frequency of (100, 400, or 1000) Hz	1701
52141C	Additional Measurement at One of the Above Frequencies	272
52150C	Physical Tests for Three-Terminal Standard Capacitors with Coaxial Connectors, Large Uncertainty (0.001 pF to 10 000 pF) at a Frequency of (100, 400, or 1000) Hz	1663
52160C	Fixed Three-Terminal Standard Capacitors with Coaxial Connectors, Large Uncertainty (0.001 pF to 10 000 pF) at a Frequency of (100, 400, or 1000) Hz	1110

52161C	Additional Measurement at One of the Above Frequencies	272
52170C	Two- or Three- Terminal Mica Dielectric Standard Capacitors with Binding Post Connectors (0.001 μF to 1 μF), at a Frequency of (66, 100, 400, 1000 or 10 000) Hz	1110
52171C	Additional Measurement at One of the Above Frequencies	272
52176C	Two-Terminal Standard Capacitors with Precision High Frequency (HF) Coaxial Connectors (0.001 pF to 10 000 pF), at a Frequency of 1000 Hz	At Cost
52180C	Fixed Standard Inductors (0.00005 H to 10 H), at a Frequency of (100, 400, 1000, or 10 000) Hz	1110
52181C	Additional Measurement at One of the Above Frequencies	1070
52190S	Special LF Inductance Measurements, by Prearrangement	At Cost

B.2 High-Frequency Standard Capacitors and Inductors

Technical Contact:	Telephone:	Email:	Mailing Address:
Ronald A. Ginley	(303) 497-3634	rginley@boulder.nist.gov	NIST
			325 Broadway, MC 818.01
			Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
52210S	Two-Terminal Low-Loss Standard Capacitors–10 kHz to 250 MHz; 1 pF to 20 pF	At Cost
52211S	Two-Terminal Low-Loss Standard Capacitors (High Accuracy)–10 kHz to 30 MHz, (50, 100, 200, 500 and 1000) pF	1394
52221C	Three-Terminal Low-Loss Standard Capacitors (High Accuracy)–10 kHz to 10 MHz, (10 ⁻² , 10 ⁻¹ , 1, 10, 10 ² and 10 ³) pF	1544
52310S	Two-Terminal, High-Q Standard Inductors (10 ⁻² µH to 1 H)	1544

B.3 Power-Frequency Capacitors

Technical Contacts: Telephone: Email: Mailing Address:

Gerald J. FitzPatrick (301) 975-8922 gfitzpatrick@nist.gov NIST

100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administive and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
52400C	Power-Frequency Capacitors	At Cost

B.4 *Q*-Standards

Technical Contact: Telephone: Email: Mailing Address:

Ronald A. Ginley (303) 497-3634 rginley@boulder.nist.gov NIST

325 Broadway, MC 818.01 Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
52710C	Inductive Q-Standards; 50 kHz to 45 MHz, 0.25 µH to 25 mH	1152
52711C	Each Additional Frequency for 52710C	180

C. Voltage Measurements

C.1 DC Voltage Measurements and Standards

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

June E. Sims (301) 975-4238 june.sims@nist.gov NIST

Yi-hua Tang (301) 975-4691 ytang@nist.gov 100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Adminstrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
53110S	Special DC Voltage Measurements, by Prearrangement	At Cost
53130C	First Saturated Standard Cell in a Group	1811
53131C	Each Additional Cell	1221
53140C	Platinum Resistance Thermometer Temperature Determination for Standard Cell Calibration	431
53150C	Unsaturated Standard Cells	1033
53160C	Tests of Solid-State Voltage Reference Standard (1 Output, 1 V to 10 V)	1614
53161C	Each Additional Output	1024
53180S	Special Handling (Equipment Pickup or Delivery)	197
53190S	Special Handling (Cleaning, Minor Repair, Return Service Charge)	397

Fees are subject to change without notice.

C.2 AC Voltage Measurements

Technical Contacts: Telephone: Email: Mailing Address:

Mark E. Parker (301) 975-2413 mparker@nist.gov NIST

Bryan C. Waltrip (301) 975-2438 bwaltrip@nist.gov 100 Bureau Drive, Stop 8170

Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
53200S	Special Tests of High-Accuracy Digital Multimeters, Multifunction Calibrators, by Prearrangement	At Cost
53201S	Special Tests of Low-Voltage AC-DC Transfer Standards, by Prearrangement	At Cost
53202S	Special 25-Point Test of Digital Multimeters (DMMs), by Prearrangement	1472
53203S	Each Additional DMM Test Point for 53202S	At Cost

C.3 AC-DC Thermal Voltage and Current Converters (to 1 MHz)

Technical Contacts: Telephone: Email: Mailing Address:

Joseph R. Kinard (301) 975-4250 jkinard@nist.gov NIST

Thomas E. Lipe (301) 975-4251 tlipe@nist.gov Building 220, Room B146
100 Bureau Drive, Stop 8170
Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
53310S	Special AC-DC Measurement Services, by Prearrangement	At Cost
53350C	Set-up Charge (No Test Points Included) for a Standard or Standards Set for AC-DC Difference (Voltage or Current)	1609
53351C	First Point for Each Applied Voltage or Current Range	757
53352C	Additional Points for Each Applied Voltage and Current Level (Additional Frequency/Voltage or Frequency/Current Points)	53

Fees are subject to change without notice.

D. Precision Ratio Measurements

D.1 Inductive Dividers

Technical Contact: Telephone: Email: Mailing Address:

Scott Shields (301) 975-4232 scott.shields@nist.gov NIST

100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
54110S	Special Ratio Measurements and Tests of Inductive Voltage Dividers, by Prearrangement	At Cost
54120C	Inductive Voltage Dividers – (Single Frequency, Voltage to be Specified, Each Setting of 3 Most Significant Dials)	3604
54121C	Additional Frequency Points	3604
54130C	Inductive Voltage Dividers – (Single Frequency, Voltage to be Specified, Each Setting of Most Significant Dial Only)	2272
54131C	Additional Frequency Points	2272

D.2 Resistive Dividers

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Gerald J. FitzPatrick (301) 975-8922 gfitzpatrick@nist.gov NIST

100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
54210C	Resistor and Resistive Dividers, Total Resistance or Voltage Ratio, Two Direct Voltage Levels Between 10 kV and 150 kV	2550
54211S	Special Tests of Resistor and Resistive Dividers at Direct Voltage Levels, by Prearrangement	At Cost
54213S	Special Tests of Resistor and Resistive Dividers at 60 Hz, by Prearrangement	At Cost

Fees are subject to change without notice.

D.3 Capacitive Dividers

Technical Contacts: Telephone: Email: Mailing Address:

Gerald J. FitzPatrick (301) 975-8922 gfitzpatrick@nist.gov NIST

100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
54310S	Special Test of Capacitive Dividers at 60-Hz, by Prearrangement	At Cost

D.4 Voltage and Current Transformers

Technical Contacts: Telephone: Email: Mailing Address:

Gerald J. FitzPatrick (301) 975-8922 gfitzpatrick@nist.gov NIST

Thomas L. Nelson (301) 975-2986 tnelson@nist.gov 100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
54510C	Voltage Transformer, Ratio & Phase Angle, at 60 Hz on 1 Range, 1 Secondary Voltage, 1 Burden Primary Vrms ≤ 150 kV	At Cost
54520C	Current Transformer, Ratio & Phase Angle, 1 Range at 1 Frequency, 1 Burden, Secondary Currents (0.5, 1, 2, 3, 4, 5) A, Primary Current Not Over 12 000 A	3394
54521C	Current Transformer, Ratio & Phase Angle, 1 Secondary Current, Additional Combination of Range, Frequency, and Burden, Primary Current Not Over 12 000 A	328
54522C	Current Transformer, Ratio & Phase at Each Additional Secondary Current, Same Combination of Range, Frequency, and Burden as 54520C or 54521C	263
54600S	Special Tests of Dividers and Transformers, by Prearrangement	At Cost

Fees are subject to change without notice.

E. Phase Meters and Standards and VOR Measurements

Technical Contacts: Telephone: Email: Mailing Address:

Mark E. Parker (301) 975-2413 mparker@nist.gov NIST

Bryan C. Waltrip (301) 975-2438 bwaltrip@nist.gov 100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics:

Denise D. Prather (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
55110S	Special Tests of Phase Standards and Related Instruments, by Prearrangement	At Cost
55120C	Phase Meters – One Combination of Input Voltages (0.5 V to 120 V) at One Frequency (2 Hz to 100 kHz) – the Input Voltage Ratio Shall Not Exceed 10	1283
55121C	Phase Meters – Each Additional Combination of Input Voltages (0.5 V to 120 V) at the Same or at a Different Frequency (2 Hz to 100 kHz) – the Input Voltage Ratio Shall Not Exceed 10	419
55130C	Phase Meters – One Additional Combination of One Input Voltage (0.5 V to 120 V) and One Input Current (1 A to 5 A) at One Frequency (2 Hz to 4 kHz)	1660
55131C	Phase Meters – Each Additional Combination of One Input Voltage (0.5 V to 120 V) and One Input Current (0.5 A to 5 A)	472

55140C	Phase Meters – One Input Voltage (120 V to 240 V) and Another Input Voltage (120 V to 240 V) at One Frequency (2 Hz to 5 kHz)	1660
55141C	Phase Meters – Each Additional Combination of One Input Voltage (120 V to 240 V) and Another Input Voltage (120 V to 240 V) at the Same or at a Different Frequency (2 Hz to 5 kHz)	472

F. Power and Energy Measurements, Low-Frequency

Technical Contacts:	Telephone:	Email:	<u>Mailing Address:</u>
Thomas L. Nelson	(301) 975-2986	tnelson@nist.gov	NIST
Gerald J. FitzPatrick	(301) 975-8922	gfitzpatrick@nist.gov	100 Bureau Drive, Stop 8170
			Gaithersburg, MD 20899-8170

Administrative and Logistics: Denise D. Prather (301) 975-4221 dprather@nist.gov

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
56110S	Special Test of AC-DC Wattmeters, by Prearrangement	At Cost
56200C	Watt, Watthour, Var, Varhour Meter, Initial Two Determinations of Same Meter at 60 Hz	3071
56201C	Each Additional Determination, Same Meter at 50 Hz	195
56202C	Initial Two Determinations of One or Two Meters Run Simultaneously with the First (56200C)	2778
56210M	Measurement Assurance Program for Watthour Meters	3995
56220S	Special Tests of Watthour Meter with Pulse Output; 120 Volts, 5 Amperes, 60 Hz at 0.5 Lag, Unity and 0.5 Lead Power Factors	1281

G. RF, Microwave and Millimeter-Wave Measurements

G.1 Thermistor Detectors

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Ronald A. Ginley (303) 497-3634 rginley@boulder.nist.gov NIST

Thomas P. Crowley (303) 497-4133 crowley@boulder.nist.gov 325 Broadway, MC 818.01 Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Service ID Number	Description of Services			
The follow	The following tests are for 50 Ω thermistor and thermoelectric detectors with coaxial connectors.			
61100S	Measurement setup charge (applies to all coaxial power measurements—one setup charge for multiple detectors with the same connectors and frequencies ¹)	1489		
61110S	Coaxial Detectors in the Frequency Range from 0.1 MHz to 10 MHz	1739		
61120S	Coaxial Detectors at user Selected Frequencies in the appropriate Frequency Range for the Connector Type ² . Up to 20 Frequency Points	1891		
61121S	Coaxial Detectors at user Selected Frequencies in the appropriate Frequency Range for the Connector Type ² . 20 to 40 Frequency Points	2081		
61122S	Coaxial Detectors at user Selected Frequencies in the appropriate Frequency Range for the Connector Type ² . 40 to 120 Frequency Points	2269		
61123S	Coaxial Detectors at user Selected Frequencies in the appropriate Frequency Range for the Connector Type ² . More than 120 Frequency Points	2648		
61137C	NIST Model CN Coaxial Detectors at 21 Frequencies within the Frequency Range of 50 MHz to 18 GHz	3767		
61138C	NIST Model CN Coaxial Detectors at Single Customer Selected Frequency within the Frequency Range of 50 MHz to 18 GHz	39		
	The following tests are for thermistor detectors with waveguide flanges.			
61140S	Measurement setup charge (applies to all waveguide power measurements EXCEPT WR15—one charge for multiple detectors with the same connectors and frequencies ¹)	2841		
61141S	Measurement setup charge (applies to all WR15 waveguide power measurements—one charge for multiple detectors with the same connectors and frequencies ¹)	2133		
61142S	Rectangular Waveguide Detectors with WR90 Flanges ²	1965		
61143S	Rectangular Waveguide Detectors with WR62 Flanges ²	1965		
61144S	Rectangular Waveguide Detectors with WR42 Flanges ²	1965		
61145S	Rectangular Waveguide Detectors with WR28 Flanges ²	2041		

61146S	Rectangular Waveguide Detectors with WR22 Flanges ²	2257
61147S	Rectangular Waveguide Detectors with WR15 Flanges ²	3544
61148S	Rectangular Waveguide Detectors with WR10 Flanges ²	4095
Miscellaneous Tests		
61190S	Special Microwave and RF Power Measurement Services, by Prearrangement	At Cost

¹ Only one setup charge is necessary for multiple detectors sent in at the same time with the same connector type and measurement frequencies.

G.2 Scattering Parameters of Passive One and Two-Port Devices

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Ronald A. Ginley (303) 497-3634 rginley@boulder.nist.gov NIST

325 Broadway, MC 818.01 Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Service ID Number	Description of Services	Fee (\$)		
	The following tests are for two-port 50 Ω devices with coaxial connectors			
61200S	Measurement setup charge for all Coaxial Two-port measurements with GR900, GC7 or Type-N connectors, frequencies of 0.05 GHz and 0.1 GHz—one setup charge for multiple devices with the same connectors and frequencies ¹	393		
61201S	Measurement setup charge for all Coaxial Two-port measurements with GR900, GC7 or Type-N connectors, frequencies at or above 0.2 GHz—one setup charge for multiple devices with the same connectors and frequencies ¹	1595		
61202C	Measurement setup charge for all Coaxial Two-port measurements with 3.5 mm, 2.92 mm, or 2.4 mm connectors and frequencies of 0.05 GHz and 0.1 GHz—one setup charge for multiple devices with the same connectors and frequencies ¹	393		
61203S	Measurement setup charge for all Coaxial Two-port measurements with 3.5 mm, 2.92 mm, or 2.4 mm connectors and frequencies at or above 0.2 GHz—one setup charge for multiple devices with the same connectors and frequencies ¹	2068		
61211S	Fixed Two-port Devices with GR900, GPC7 or Type-N connectors ²	2395		
61212S	Variable Two-port Devices with GR900, GPC7 or Type-N connectors ²	2395		
61213S	Fixed Two-port Devices with 3.5 mm, 2.92 mm, or 2.4 mm connectors ²	1833		
61214S	Variable Two-port Devices with 3.5 mm, 2.92 mm, or 2.4 mm connectors ²	1915		

² Measurement Frequencies

	The following tests are for two-port devices with waveguide flanges.	
61220S	Measurement setup charge for all Waveguide Two-port measurements EXCEPT WR15—one setup charge for multiple devices with the same connectors and frequencies ¹	2032
61221S	Measurement setup charge for all WR15 Waveguide Two-port measurements—one setup charge for multiple devices with the same connectors and frequencies ¹	2333
61222S	Additional Measurement setup charge for all VARIABLE Waveguide Two-port measurements— one setup charge for multiple devices with the same connectors and frequencies ¹	429
61223S	Fixed and Variable Rectangular Waveguide Two-port Devices with WR90 Flanges ²	1876
61224S	Fixed and Variable Rectangular Waveguide Two-port Devices with WR62 Flanges ²	1871
61225S	Fixed and Variable Rectangular Waveguide Two-port Devices with WR42 Flanges ²	1953
61226S	Fixed and Variable Rectangular Waveguide Two-port Devices with WR28 Flanges ²	1953
61227S	Fixed and Variable Rectangular Waveguide Two-port Devices with WR22 Flanges ²	1953
61228S	Fixed and Variable Rectangular Waveguide Two-port Devices with WR15 Flanges ²	3734
61231S	Fixed and Variable Rectangular Waveguide Two-port Devices with WR10 Flanges ²	3816
	The following tests are for one-port 50 Ω devices with coaxial connectors	
61250S	Measurement setup charge for all Coaxial One-port measurements with GR900, GC7 or Type-N connectors, frequencies of 0.05 GHz and 0.1 GHz—one setup charge for multiple devices with the same connectors and frequencies ¹	504
61251S	Measurement setup charge for all Coaxial One-port measurements with GR900, GC7 or Type-N connectors, frequencies at or above 0.2 GHz—one setup charge for multiple devices with the same connectors and frequencies ¹	1653
61252S	Measurement setup charge for all Coaxial One-port measurements with 3.5 mm, 2.92 mm, or 2.4 mm connectors and frequencies of 0.05 GHz and 0.1 GHz — one setup charge for multiple devices with the same connectors and frequencies ¹	477
61253S	Measurement setup charge for all Coaxial One-port measurements with 3.5 mm, 2.92 mm, or 2.4 mm connectors and frequencies at or above 0.2 GHz—one setup charge for multiple devices with the same connectors and frequencies	2094
61261S	Fixed One-port Devices with GR900, GPC7 or Type-N connectors ²	2091
61262S	Variable One-port Devices with GR900, GPC7 or Type-N connectors ²	2091
61263S	Fixed One-port Devices with 3.5 mm, 2.92 mm, or 2.4 mm connectors ²	1712
61264S	Variable One-port Devices with 3.5 mm, 2.92 mm, or 2.4 mm connectors ²	1876

	The following tests are for one-port devices with waveguide flanges.				
61270S	Measurement setup charge for all Waveguide One-port measurements —one setup charge for multiple devices with the same connectors and frequencies ¹	2114			
61271S	Additional Measurement setup charge for all VARIABLE Waveguide One-port measurements —one setup charge for multiple devices with the same connectors and frequencies ¹	422			
61272S	Fixed and Variable Rectangular Waveguide One-port Devices with WR90 Flanges ²	1712			
61273S	Fixed and Variable Rectangular Waveguide One-port Devices with WR62 Flanges ²	1789			
61274S	Fixed and Variable Rectangular Waveguide One-port Devices with WR42 Flanges ²	1789			
61275S	Fixed and Variable Rectangular Waveguide One-port Devices with WR28 Flanges ²	1789			
61276S	Fixed and Variable Rectangular Waveguide One-port Devices with WR22 Flanges ²	1789			
61277S	Fixed and Variable Rectangular Waveguide One-port Devices with WR15 Flanges ²	3683			
61278S	Fixed and Variable Rectangular Waveguide One-port Devices with WR10 Flanges ²	3683			
	Miscellaneous Tests				
61290S	Special Microwave and RF Power Measurement Services, by Prearrangement	At Cost			

¹ Only one setup charge is necessary for multiple detectors sent in at the same time with the same connector type and measurement frequencies

G.3 High-Accuracy Attenuation Measurements

Technical Contacts:	Telephone:	<u>Email:</u>	Mailing Address:
Ronald A. Ginley	(303) 497-3634	rginley@boulder.nist.gov	NIST
Jeff Jargon	(303) 497-3516	jargon@boulder.nist.gov	325 Broadway, MC 818.01
			Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Service ID Number	Description of Services	Fee (\$)
61300S	High-Accuracy Attenuation: Set Up Charge, per order	At Cost

² Measurement Frequencies

61330S	Attenuation Measurements of Three-Port and Two-Port Devices at 1.25 MHz, 0 dB and 6 dB	At Cost
61350C	Coaxial Fixed and Variable Phase Shifters; Characteristics Phase Shift Difference; Precision Connectors; Measured at 30 MHz, Range 0° to 360°	At Cost

G.4 Thermal Noise Measurements

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

David Walker (303) 497-5490 dwalker@boulder.nist.gov NIST

James Randa (303) 497-3150 randa@boulder.nist.gov 325 Broadway, MC 818.01

Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Service ID Number	Freq.	Connector Type	Device Requirements/Service	Fee (\$)
61410S	30 MHz 60 MHz Set Up Charge	-	Temperature < 15 000 K (ENR < 17 dB) VSWR < 1.2	2380
61420S	1.0 GHz to 12.4 GHz Continuous Frequencies	Coaxial 14 mm (1 to 4 GHz) GPC 7 N Precision (PIN) GPC 3.5 (PIN) GPC 2.4 (PIN) (8 to 12.4 GHz)	Temperature < 15 000 K (ENR < 17 dB) Reflection Coefficient < 0.2	3765
	Set Up Charge,	, per order		3273
61425S	12.4 GHz to 18.0 GHz Continuous Frequencies Set Up Charge	Coaxial GPC 7 N Precision (PIN) GPC 3.5 (PIN) GPC 2.4 (PIN)	Temperature < 15 000 K (ENR < 17 dB) Reflection Coefficient < 0.2	371 4405 2491
	Per Frequency			2491

61430S		Coaxial	Temperature < 15 000 K	
014305	18.0 GHz to	GPC 3.5 (PIN)	(ENR < 17 dB)	
	26 GHz	GPC 2.4 (PIN)	Reflection Coefficient < 0.2	
	Continuous	, ,		
	Frequencies			1200
	Set Up Charge,	per order		4299 2396
	Per Frequency	Coaxial	Tomporoture < 15 000 V	2390
61435S	26.5 GHZ to	GPC 2.4 (PIN)	Temperature < 15 000 K (ENR < 17 dB)	
	40 GHz	GI C 2.4 (I IIV)	Reflection Coefficient < 0.2	
	Continuous		Reflection Coefficient < 0.2	
	Frequencies			
	Set Up Charge,	per order	- I	6167
	Per Frequency	Pro serve		3503
61450S	1	Waveguide	Temperature <15 000 K	
014505	8.2 GHz to	WR 90	(ENR < 17 dB)	
	12.4 GHz		Reflection Coefficient < 0.2	
	Continuous			
	Frequencies			
	Set Up Charge,	per order		3381
	Per Frequency	T		363
61455S		Waveguide	Temperature < 15 000 K	
	12.4 GHz to	WR 62	(ENR < 17 dB)	
	18.0 GHz		Reflection Coefficient < 0.2	
	Continuous Frequencies			
	Set Up Charge,	ner order		4299
	Per Frequency	per order		2313
61460S	18.0 GHz to	Waveguide	Temperature < 15 000 K	2010
014003	26.0 GHz	WR 42	(ENR < 17 dB)	
	Continuous	WK 42	Reflection Coefficient < 0.2	
	Frequencies		Reflection Coefficient < 0.2	
	Set Up Charge,	per order		4299
	Per Frequency	Т		2313
61465S	26.5.011.4	Waveguide	Temperature < 15 000 K	
	26.5 GHz to 40.0 GHz	WR 28	(ENR < 17 dB)	
	Continuous		Reflection Coefficient < 0.2	
	Frequencies			
	Set Up Charge,	per order	•	5596
	Per Frequency			2611
61470S		Waveguide	Temperature <15 000 K	
	33 GHz to	WR 22	(ENR < 17 dB)	
	50 GHz Continuous		Reflection Coefficient <0.2	
	Frequencies			
	Set Up Charge,	per order	•	5976
	Per Frequency			3289

61475S	50 GHz to 65 GHz Continuous Frequencies	Waveguide WR 15	Temperature < 15 000 K (ENR , 17 dB) Reflection Coefficient < 0.2	
	Set Up Charge,	per order		5976
	Per Frequency			4254
61495S	Special Noise To	emperature Measur	rements, by Prearrangement	At Cost

G.5 Microwave Dielectric and Magnetic Material Measurements

Technical Contact: Telephone: Email: Mailing Address:

James R. Baker-Jarvis (303) 497-5621 jjarvis@boulder.nist.gov NIST

325 Broadway, MC 818.01 Boulder, CO 80305-3328

Administrative and Logistics: Puanani L. DeLara

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Service ID Number	Description of Services	Fee (\$)
61620S	Special Tests for Dielectric and Magnetic Materials 1 kHz to 60 GHz	At Cost
61640S	Special Consulting and Advisory Services for Dielectric and Magnetic Materials, by Prearrangement	At Cost

H. Electromagnetic Field Strength and Antenna Measurements

H.1 Microwave Antenna Parameter Measurements

Technical Contacts: Telephone: Email: Mailing Address: Perry F. Wilson (303) 497-3406 pfw@boulder.nist.gov **NIST** (63100S-63400S) 325 Broadway, MC 818.02 Katherine MacReynolds (303) 497-3471 mreynold@boulder.nist.gov Boulder, CO 80305-3328

(63100S)

Michael H. Francis (303) 497-5873 francis@boulder.nist.gov

(63200S)

Administrative and Logistics:

(303) 497-3753 calibration@boulder.nist.gov Puanani L. Delara

Fax: (303) 497-3970

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
63100S	Gain and Polarization Calibrations of Standard Antennas Using Extrapolation Range	At Cost
63200S	Measurement of Pattern, Gain, and Polarization of Arbitrary Antennas Using Near-Field Scanning Techniques	At Cost
63300S	Special Test Service for Calibration of Probes Used with Near-Field Scanning Facilities	At Cost
63400S	Special Consulting, Advisory, and Other Services	At Cost

H.2 Field Strength Parameter Measurements

Technical Contacts: Email: **Mailing Address: Telephone:**

Dennis G. Camell (303) 497-3214 camell@boulder.nist.gov NIST

Perry F. Wilson (303) 497-3406 pfw@boulder.nist.gov 325 Broadway, MC 818.02 Boulder, CO 80305-3328

Administrative and Logistics:

Puanani L. DeLara (303) 497-3753 calibration@boulder.nist.gov

Fax: (303) 497-3970

Service ID Number	Description of Services	Fee (\$)
64100S	Special Test Services for Antenna/Field Strength/Measurement, Using the Transverse Electromagnetic (TEM) Cell Method (10 kHz to 300 MHz)	At Cost
64200S	Special Test Services for Antenna/Field Strength/Measurements, Utilizing the Open Area Test Site and Standard Antenna Method	At Cost
64300S	Special Test Services for Antenna/Field Strength/Reflectivity Measurements, Utilizing the Anechoic Chamber and Standard Field Method	At Cost

I. Pulse Waveform Measurements

<u>Technical Contacts:</u> <u>Telephone:</u> <u>Email:</u> <u>Mailing Address:</u>

Paul Hale (65100S, 65200S and (303) 497-5367 hale@boulder.nist.gov NIST

65400S) 325 Broadway, MC 815.01

Boulder, CO 80305-3328

David I. Bergman (65250S only) (301) 975-4464 dbergman@nist.gov NIST

100 Bureau Drive, Stop 8170 Gaithersburg, MD 20899-8170

Administrative and Logistics (65100S, 65200S and 65400S):

John Lomax (303) 497-4285

FAX: (303) 497-4286

Administrative and Logistics (65250S only):

Denise D. Prather (65250S only) (301) 975-4221 dprather@nist.gov

Service ID Number	Description of Services	Fee (\$)
65100S	Impulse Spectrum Amplitude (50 Ω)	At Cost
65200S	Fast Repetitive Pulse Transition Parameters (50 Ω)	At Cost
65250S	Fast Repetitive Pulse Settling Parameters (50 Ω)	At Cost
65400S	Pulse Time Delay Interval	At Cost

CHAPTER 10

TIME AND FREQUENCY MEASUREMENTS

A. Broadcast and Measurement Services

Technical Contacts:	Telephone:	<u>Email:</u>	Mailing Address:
Michael A. Lombardi (Frequency)	(303) 497-3212	lombardi@boulder.nist.gov	NIST
Marc A. Weiss (Time)	(303) 497-3261	mweiss@boulder.nist.gov	325 Broadway, MC 847.40
John Lowe	(303) 497-5453	lowe@boulder.nist.gov	Boulder, CO 80305-3328

Administrative and Logistics: Trudi Peppler (303) 497-3338 tpeppler@boulder.nist.gov

Fax: (303) 497-6461

Do not ship instruments or standards to the mailing address. Contact the staff for the shipping address.

Service ID Number	Description of Services	Fee (\$)
	Broadcast Services (WWW, WWVH, WWVB, GOES, ACTS	S and NTS)
76100S	Frequency Measurement Service (Frequency Delivered to User's Site)	Initial One-Time Fee \$1,500 Monthly Charge \$500
76110S	Global Time Service (Frequency and Time delivered to User's Site)	\$15,000 per year
76120S	Characterization of Global Positioning System (GPS) Satellite Receivers	At Cost

B. Calibration and Characterization of Oscillators and Amplifiers

<u>Technical Contacts:</u>	<u>Telephone:</u>	<u>Email:</u>	<u> Mailing Address:</u>
David Howe	(303) 497-3277	dhowe@boulder.nist.gov	NIST
James E. Gray	(303) 497-3209	jgray@boulder.nist.gov	325 Broadway, MC 847
			Boulder, CO 80305-3328

Administrative and Logistics: Trudi Peppler (303) 497-3338 tpeppler@boulder.nist.gov

Fax: (303) 497-6461

Service ID Number	Description of Services	Fee (\$)
77100C	Oscillator Frequency Calibration	At Cost
77110C	Characterization of Atomic Frequency Standards	At Cost
77120C	Characterization of Oscillators: Time Domain	At Cost

77130C	Characterization of Oscillators and Amplifiers: Phase Noise in the Frequency Domain	At Cost
77131C	Characterization of Oscillators and Amplifiers: Amplitude Noise in the Frequency Domain	At Cost

C. Test of PM/AM Noise Measurement Systems

Technical Contact: Telephone: Email: Mailing Address:

David Howe (303) 497-3277 dhowe@boulder.nist.gov NIST

325 Broadway, MC 847.30 Boulder, CO 80305-3328

Administrative and Logistics:

Trudi Peppler (303) 497-3338 tpeppler@boulder.nist.gov

Fax: (303) 497-6461

Service ID Number	Description of Services	Fee (\$)
77135C	Tests of RF PM/AM Noise Measurement Systems: On-Site Tests	At Cost
77136C	Tests of Microwave PM/AM Noise Measurement Systems: On-Site Tests	At Cost
77140S	Special Time/Frequency Measurements: Oscillators and Other Components	At Cost

CHAPTER 11 SEMINARS

The following announcements concern notification of changes in services and information about future NIST Measurement Seminars. General policy question regarding NIST measurement services should be referred to the Calibration Program.

NIST MEASUREMENT SEMINARS

NIST holds seminars and workshops that provide advice and assistance on measurements and calibrations. This affords laboratories outside NIST and opportunity to learn how to make measurements consistent with national standards which NIST maintains. Participation is open to a limited number of people who have the appropriate education, work experience, and current profession in measurement and standards laboratory activities.

Each seminar lasts from one to five days and is devoted to lectures, group discussions, and laboratory demonstrations. A course may be cancelled if registration is insufficient. However, in the past, requests for enrollment have nearly always exceeded the numbers that could be accommodated.

Acceptance letters will be mailed no later than 4 weeks prior to the scheduled date of the course. Detailed information on schedules and housing will be included. Those accepted will be expected to study the assigned reading material before coming to the course and should be prepared to discuss their own experiences with related problems.

See the Weights and Measures Program web site www.ts.nist.gov/ts/htdocs/230/235/calen103.htm for the NCWM Calendar of Events for other training not listed here.

NIST offers conferences and workshops throughout the year. To see the latest listing go to www.nist.gov/public_affairs/confpage/conffutr.htm for Upcoming NIST Conferences.

ALPHABETICAL CROSS-INDEX

<u>Item</u> <u>Service ID Numbers</u>

\boldsymbol{A}

Absolute pressure transducers	30010C-30011C
Accelerometers	
Ac-dc thermal converters (to 1MHz)	53350C-53352C
Ac-dc watthour or varhour meters	56110S-56210M
Ac voltage, high accuracy	53200S-53203S
Acoustic devices	25060S
Acoustic measurements	25010C-25070S
Activation detector dosimetry	44070C-44090C
Air navigation aids	55210S-55230S
Air-speed indicators	19010C-19040S
Alpha-particle-emitting solid sources	43030C-43050C
American Petroleum Institute gages	12010C
Anechoic chambers	25010C-25070S
Anemometers	19010C-19040S
Angle gage blocks	14010C
Angular measurements	14010C-14050S
Antenna parameter measurements, microwave	63100S-63400S
Attenuators, rf and microwave	61210S-61250S
Audio-frequency phase meter	55120C-55141C
\boldsymbol{B}	
D	
D. H. J.	110500
Ball plates	
Balls	
Barometers	
Beckman thermometers	
Bell provers	
Beta-particle applications	
Beta-particle emitting sources	
DI 11 1	
Blackbody sources	
Bolometer units, rf and microwave	61110S-61138S
\boldsymbol{C}	
Calorimeter thermometers	31170S
Capacitance and inductance measurements, high frequency	
Capacitance and inductance measurements, low frequency	
Capacitive voltage dividers	
Capacitors, dielectric	
Capacitors, high frequency	
Capacitors, low frequency	
Capacitors, power frequency	
Coaxial attenuators	
Coaxial thermistor detectors	
Color measuring instruments	
•	
Coordinate massuring mechines	
Coordinate measuring machines.	110303, 120403,

	15010C-15040C
-software algorithms	
-socketed ball bars	
-probe performance spheres	A ,
Cryogenic flow measurements	
Cryogenic resistance thermometers	
Current and voltage transformers	
D	
Dc resistance measurements	51100S-51163C
Dc voltage measurements	53110S-53190S
Deadweight piston gages	
Density measurements, liquids	
Detector standards, windowed photodiode	
Detector standards, windowless photodiode	
Detectors, IR	
Detectors, near ultraviolet to soft x-ray region	
Detectors, radiometric	39080S
Detectors, spectroradiometric	39071C-39090S
Detectors, used with lasers	42110C-42180S
Deuterium lamps	39050C
Dew-point hygrometer	36010C-36020C
Diameter measurements	11010S-11060S
Dielectric materials	61620S
Differential pressure transducers	30020C-30021C
Digital multimeters, low frequency	53200S-53203S
Dimensional metrology	10010C-14050S
Dividers, inductive	54110S-54131C
Dividers, resistive	54210C-54213S
Dosimeters, electron beam	48010M-48020S
Dosimeters, high dose	49010C-49050S
Dosimeters, neutron	44100S
Dosimeters, radiochromic	49010C-49030C
Dosimeters, x-ray, gamma-ray, and electron	
Dosimetry, neutron	
$oldsymbol{E}$	
Earphones	25070S
Electromagnetic field-strength parameter measurements	
Electrometers	
Electron beam dosimetry	
End standards	
Energy and power measurements low frequency	56110S-56210M

\boldsymbol{F}

Ferrous-ferric dosimeters.	
Fiber optic power meters	42130C-42180S
Field strength measurements, electromagnetic	64100S-64300S
Filament lamps, ribbon	35050C-35060C,
	39010C-39030C
Filters, spectral transmittance	38010C-38040C
Fixed-point devices, thermometric	
Flashing-light photometer	
Flats, optical reference	
Flow measurements, cryogenic	
Flowrate meters, liquid and gas	
Flux standards, luminous	
Force transducers	
Frequency and time measurement services	
Frequency domain	
	,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,
G	
Gage blocks	10010C
Gages, API	
Gages, ionization	
Gages, low pressure	
Gages, piston	
Gages, plug	
Gages, plug and ring	
Gages, pressure	
Gages, ring	
Gages, step	
Gages, two dimensional	
Gages, vacuum	
Gamma-ray dosimeters	
Gamma-ray emitting sources	
Gas flowmeters	
Gear measuring wires	
Germanium resistance thermometers	
Global time service	
Global positioning system satellite receivers	
Grid plates	11050S
H	
Heat flux sensors	35100S-35102C
Helium permeation leaks	30060S-30062S
High dose dosimetry	
High-frequency standard capacitors and inductors	
High-frequency standard resistors	
High-vacuum gages	
High-voltage resistors	
Humidity measurements	
Hydrocarbon liquid flow meters	
Hydrometers	
Hygrometers	
<i>y</i>	

I

Illuminance meters	. 37090S
Impedance measurements	. 52110S
Impulse spectrum amplitude	
Impulse response of coaxial networks	
Incandescent lamps	
Indexing tables	
Inductance and capacitance measurements, low frequency	
Inductance and capacitance measurements, high frequency	
Inductive voltage dividers	
Industrial grade thermometers	
Intensity, luminous	
Interferometry	
Internal diameter standards	
Ionization chambers.	
Ionization gages	
Irradiance detectors	
Irradiance sources	
Traditate sources	.370400 370300
$oldsymbol{L}$	
Laboratory thermometers	. 31010C-31180S
Lamps, deuterium arc	
Lamps, incandescent	
Lamps, quartz halogen	
Lamps, ribbon filament	
Lamps, 1000n manene	
Laser frequency/wavelength	
Laser power and energy measurements	
Leak artifacts	
Length and diameter	
Length measurements	
Length standards	
Light emitting diodes (LEDs)	
Line standards	
Linearity of optical fiber power meters	
Liquid flow meters	
Liquid-in-glass thermometers.	
Load cells	
Luminous flux	
Luminous intensity of LEDs	
Luminous intensity of LEDs	. 3/1303
M	
Mammography x-ray measuring instruments	. 46010C-46050S
Manometers	
Mass measurements	
MEASUREMENT ASSURANCE PROGRAMS (MAPs)	C1110N4
Do resistance.	
Dose interpretation of ferrous-ferric dosimeters	
Laser power and energy	
Platinum resistance thermometry	
Watthour meters	. 5021UM

	21010G 21100G
Mercury-in-glass thermometers	
Meteorological airspeed instrumentation	
Meters, laser power and energy	
Meters, optical fiber power	
Meters, phase	55110S-55141C
Microphones, pressure and free-field response	25010C-25050C
Micropotentiometers, low voltage	53201S
Micropotentiometers, rf	
Microwave antenna parameter measurements	
Molecular drag gages	
N	
Neutron dosimeters	44100S
Neutron sources	
Noise temperature measurements, coaxial and waveguide	
Non-mercurial barometers and manometers	
Non-inercurial varonieters and manometers	29033C
o	
Oil gaging and surveying tapes	10030C
Opal glass luminance standards	
Optical fiber power meters	
Optical flats	
Optical polygons	
Optical pyrometers	
Optical reference planes	
Optoelectronic Frequency Response	
Oscillator characterization	
Penetrameters, x-ray	77100C-77140S 46040S
Oscillator characterization	77100C-77140S 46040S
Penetrameters, x-ray	77100C-77140S 46040S 44060C
Penetrameters, x-ray Personnel protection instrumentation	46040S 44060C 55110S-55141C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave	46040S 44060C 55110S-55141C 61295S-61297S
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response.	46040S 44060C 55110S-55141C 61295S-61297S 39070C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared	46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless Photographic step tablets	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5)
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless Photographic step tablets Photometers	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless Photographic step tablets Photometers Photometers Photometric tests	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photographic step tablets Photometers Photometers Photometric tests Photometry	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S 37010C-37110S
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photographic step tablets Photometers Photometers Photometric tests Photometry Photon Beams	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S 37010C-37110S 49010C-49050S
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photographic step tablets Photometers Photometers Photometry Photon Beams Pickup sensitivity	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S 37010C-37110S 49010C-49050S 24010C-24030C, 24050C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photographic step tablets Photometers Photometers Photometric tests Photometry Photon Beams Pickup sensitivity Piston gages, controlled clearance	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S 37010C-37110S 49010C-49050S 24010C-24030C, 24050C 29020C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless Photographic step tablets Photometers Photometers Photometric tests Photometry Photon Beams Pickup sensitivity Piston gages, controlled clearance Piston gages, dead weight	77100C-77140S46040S44060C55110S-55141C61295S-61297S39070C39077C40510C-40560CSee SRMs (p.5)37090S37100S37010C-37110S49010C-49050S24010C-24030C, 24050C29020C29010C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless Photographic step tablets Photometers Photometers Photometric tests Photometry Photon Beams Pickup sensitivity Piston gages, controlled clearance Piston gages, dead weight Piston gauges versus UIM	77100C-77140S46040S44060C55110S-55141C61295S-61297S39070C39077C40510C-40560CSee SRMs (p.5)37090S37100S37010C-37110S49010C-49050S24010C-24030C, 24050C29020C29010C30025C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless Photographic step tablets Photometers Photometric tests Photometric tests Photometry Photon Beams Pickup sensitivity Piston gages, controlled clearance Piston gages, dead weight Piston gauges versus UIM Pitot-static tubes	77100C-77140S46040S44060C55110S-55141C61295S-61297S39070C39077C40510C-40560CSee SRMs (p.5)37090S37100S3710C-37110S49010C-49050S24010C-24030C, 24050C29020C29010C30025C19010C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiode, ultraviolet to near-infrared Photodiodes, windowed and windowless Photographic step tablets Photometers Photometric tests Photometry Photon Beams Pickup sensitivity Piston gages, controlled clearance Piston gages, dead weight Piston gauges versus UIM Pitot-static tubes Platinum resistance thermometers	77100C-77140S46040S44060C55110S-55141C61295S-61297S39070C39077C40510C-40560CSee SRMs (p.5)37090S37100S3710C-37110S49010C-49050S24010C-24030C, 24050C29020C29010C30025C19010C33010C-33310C
Penetrameters, x-ray	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S 37100S 37010C-37110S 49010C-49050S 24010C-24030C, 24050C 29020C 29010C 30025C 19010C 33010C-33310C 12010C-12020C
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiodes, windowed and windowless Photographic step tablets Photometers Photometric tests Photometry Photon Beams Pickup sensitivity Piston gages, controlled clearance Piston gages, dead weight. Piston gauges versus UIM Pitot-static tubes Platinum resistance thermometers Plug and ring gages Plug gages	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S 37010C-37110S 49010C-49050S 24010C-24030C, 24050C 29020C 29010C 30025C 19010C 33010C-33310C 12010C-12020C 11010S
Penetrameters, x-ray	77100C-77140S 46040S 44060C 55110S-55141C 61295S-61297S 39070C 39077C 40510C-40560C See SRMs (p.5) 37090S 37100S 37010C-37110S 49010C-49050S 24010C-24030C, 24050C 29020C 29010C 30025C 19010C 33010C-33310C 12010C-12020C 11010S
Penetrameters, x-ray Personnel protection instrumentation Phase measurements Phase shifters, rf and microwave Photodiode, spectral response Photodiodes, windowed and windowless Photographic step tablets Photometers Photometric tests Photometry Photon Beams Pickup sensitivity Piston gages, controlled clearance Piston gages, dead weight. Piston gauges versus UIM Pitot-static tubes Platinum resistance thermometers Plug and ring gages Plug gages	77100C-77140S46040S44060C55110S-55141C61295S-61297S39070C39077C40510C-40560CSee SRMs (p.5)37100S37100S37010C-37110S49010C-49050S24010C-24030C, 24050C29020C29010C30025C19010C33010C-33310C12010C-12020C11010S77135C-77136S
Penetrameters, x-ray	77100C-77140S46040S44060C55110S-55141C61295S-61297S39070C39077C40510C-40560CSee SRMs (p.5)37090S37100S37010C-37110S49010C-49050S24010C-24030C, 24050C29020C29010C30025C19010C33010C-33310C12010C-12020C11010S77135C-77136S36060C

Power and energy measurements, low frequency	56110S-56210M		
Power meters, optical fiber			
Power meters, thermoelectric			
Pressure gages and transducers.			
Pressure gages, low absolute			
Pressure gages, low differential			
Pressure gages, special tests			
Pressure measurements			
r ressure measurements.			
Proving rings			
Pulse waveform measurements			
Pyrometer indicators			
Pyrometers, optical			
Pyrometers, radiation	350/0 5 -350/2C		
Q			
~			
Q-standards	52710C-52711C		
Quartz halogen lamps			
Quartz narogen rumps	370400 370430		
ת			
R			
Radiance standard sources	39010C-39030C		
Radiance thermometry	35010C-35084C		
Radiant energy detectors			
Radiation detectors			
Radiation pyrometers	35070S		
Radiation thermometry			
Radioactive sources, alpha emitting			
Radioactive sources, beta emitting			
Radioactive sources, gamma emitting			
Radioactive sources, neutron emitting			
Radiochrometric dosimeters			
Radiometric detectors			
Radiometric measurements			
Radiometric sources			
Radionuclides			
Reflection coefficients, coaxial and waveguide terminals			
Resistance measurements, dc			
Resistance thermometers			
Resistive voltage dividers			
Resistors, high frequency			
Resistors, high-precision standard, dc			
Resistors, high voltage.			
Resistors, megohm			
Resistors for current measurements			
Ring gages			
R-meters			
Rods, surveying leveling			
Rotary and indexing tables			
Roughness calibration standards			
Roughness calibration standards	15010C-15040S		

S

Saturated standard cells			
Shock measurements			
Sieves, special tests			
Spectral irradiance			
Spectral irradiance, lamps			
Spectral radiance			
Spectral radiance, lamps			
Spectral reflectance	38060S		
Spectral response, lasers			
Spectral response, photodiode			
Spectral transmittance filters			
Spectroradiometric detectors			
Spectroradiometric sources	39010C-39060S		
Specular gloss			
Spherical diameter standards, balls	11030S		
Stage micrometer	10020C		
Standard voltage cells	53130C-53150C		
Step gages, length	11060S		
Step height measurements, surface	15030C		
Step tablet density, photographic	See SRMs (p. 5)		
Sucker rods	12010C		
Surface color	38091S		
Surface roughness standards	15040S		
Surveying and oil gaging tapes	10030C		
Surveying leveling rods			
T	100000		
Tapes			
Tapes Thermal voltage converters (to 1 MHz)	53350C-53352C		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts	53350C-53352C 61110S-61155S		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts Thermocouples and thermocouple materials Thermometers, Beckmann	53350C-53352C 61110S-61155S 32010C-32150S 31180S		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts Thermocouples and thermocouple materials Thermometers, Beckmann Thermometer, calorimetric type	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts Thermocouples and thermocouple materials Thermometers, Beckmann Thermometer, calorimetric type Thermometers, cryogenic type Thermometers, germanium resistance Thermometers, laboratory	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts Thermocouples and thermocouple materials Thermometers, Beckmann Thermometer, calorimetric type Thermometers, cryogenic type Thermometers, germanium resistance Thermometers, laboratory Thermometers, platinum resistance	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts Thermocouples and thermocouple materials Thermometers, Beckmann Thermometer, calorimetric type Thermometers, cryogenic type Thermometers, germanium resistance Thermometers, laboratory Thermometers, platinum resistance Thermometer, rhodium-iron resistance	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 31010C-31260S 33010C-33310C 33140C-33141C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5)		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts Thermocouples and thermocouple materials Thermometers, Beckmann Thermometer, calorimetric type Thermometers, cryogenic type Thermometers, germanium resistance Thermometers, laboratory Thermometers, platinum resistance Thermometer, rhodium-iron resistance Thermometric fixed-point devices Thermometry, radiance Thermometry, radiance Thread measuring wires	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S 11020C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S 11020C 12020S		
Tapes Thermal voltage converters (to 1 MHz) Thermistor mounts Thermocouples and thermocouple materials Thermometers, Beckmann Thermometer, calorimetric type Thermometers, cryogenic type Thermometers, germanium resistance Thermometers, laboratory Thermometers, platinum resistance Thermometer, rhodium-iron resistance Thermometric fixed-point devices Thermometry, radiance Thread measuring wires Threaded plug and ring gages Time and frequency measurement services	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 3360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 3360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 3360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C 30010C-30021C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 3360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C 30010C-30021C 54510C-54600S		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 3360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C 30010C-30021C 54510C-54600S 38010C-38040C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C 30010C-30021C 54510C-54600S 38010C-38040C 37080S		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C 30010C-30021C 54510C-54600S 38010C-38040C 37080S 39010C-39046C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C 30010C-30021C 54510C-54600S 38010C-38040C 37080S 39010C-39046C 35050C-35061C		
Tapes	53350C-53352C 61110S-61155S 32010C-32150S 31180S 31170S 33140C-33141C 33142C 31010C-31260S 33010C-33310C 33140C-33141C 33360S also SRMs (p. 5) 35010C-35070S 11020C 12020S 76100S-76120C 15040S 24010C-24030C, 24050C 30010C-30021C 54510C-54600S 38010C-38040C 37080S 39010C-39046C 35050C-35061C		

$oldsymbol{U}$

Ultraviolet detectors	40510C-40599S
Unsaturated standard cells	
V	
Vacuum gages	30029C-30038C, 30050S
Vacuum measurements	
Var, varhour meters	
Very-high-frequency omnidirectional measurements	
Vibration and shock measurements	
Voltage, ac	
Voltage, dc	53110S-53190S
Voltage, rf dc	
Voltage and current transformers	54510C-54600S
Voltage dividers, capacitive	54310S
Voltage dividers, inductive	54110S-54131C
Voltage dividers, resistive	54210C-54213S
Voltage reference standards, solid state	53160C-53161C
Voltmeters, high accuracy, ac	53200S
Volume standards	17010C-17040S
VOR measurements	55110S
W	
Water flow meters	18020C. 18060S
Watt, watthour meters	
Wattmeters, high power	
Wattmeters, high power	
Waveform measurements, pulse	
Waveguide attenuators	
Waveguide terminations and reflection coefficients	
Waveguide thermistor detectors	
Wavelength standards	
Wedges	
Weights (mass standards)	
Wires for thread and gear measurements.	

X

X-ray	and photographic step tables	See SRMs (p. 5)
•	dosimeters	•
	penetrameters	
•	•	